

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0161 Slanetz & Bartley Medium
Batch Number: 00003093
Expiry Date: 04.2022
Date Received: 04.2019
Date Tested: 04.2019
Sample Condition: Satisfactory
Sample Number: 356076

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP0260 with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Enterococcus faecalis NCTC 12697/ATCC 29212	82	Maroon colonies 1.0-1.5mm	Maroon colonies 1.0-1.5mm
Enterococcus faecium DSMZ 2146	110	Maroon colonies 1.0-1.5mm	Maroon colonies 1.0-1.5mm
Enterococcus faecium NCTC 12202	119	Maroon colonies 1.0-1.5mm	Maroon colonies 1.0-1.5mm

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2 .		
Selectivity	Test Result	Specification
Escherichia coli NCTC 12241/ATCC 25922	Inhibited	Inhibited
Staphylococcus aureus NCTC 12981/ATCC 25923	Inhibited	Inhibited

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0161 Slanetz & Bartley Medium
Batch Number: 00003093
Expiry Date: 04.2022
Date Received: 04.2019
Date Tested: 04.2019
Sample Condition: Satisfactory
Sample Number: 356076

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Physical	Result	Specifications
pH	7.4	7.2 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Rose

All of the results on this certificate of analysis relate only of the samples submitted.

Douglas Cameron
Technical Manager, E&O Diagnostics Ltd