

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0019 Yeast Extract Glucose Chloramp
Batch Number: 00003278
Expiry Date: 07.2022
Date Received: 07.2019
Date Tested: 07.2019
Sample Condition: Satisfactory
Sample Number: 359713

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP1191 with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Aspergillus brasiliensis NCPF 2275/ATCC 16404	120	White colonies with black pigmentation. Reverse side is pale yellow.	White colonies with black pigmentation. Reverse side is pale yellow.
Candida albicans NCPF 3179/ATCC 10231	97	White colonies	White colonies
Saccharomyces cerevisiae NCPF 3178	103	White colonies	White colonies

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2 .

Selectivity	Test Result	Specification
Escherichia coli NCTC 12241/ATCC 25922	Inhibited	Inhibited

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0019 Yeast Extract Glucose Chloramp
Batch Number: 00003278
Expiry Date: 07.2022
Date Received: 07.2019
Date Tested: 07.2019
Sample Condition: Satisfactory
Sample Number: 359713

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Physical	Result	Specifications
pH	6.8	6.6 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Pale Straw2-8 - 6-8

All of the results on this certificate of analysis relate only of the samples submitted.



Douglas Cameron
Technical Manager, E&O Diagnostics Ltd