

## SIMICON

### Quality certificate

#### Biological indicators SIMICON ST/DA-V6

Biological indicators in compliance with ISO 11138-3 for steam sterilization at 121 °C/134 °C.

Organism: *Geobacillus stearothermophilus* spores ATCC 7953

Mean population (cfu):  $1,8 \times 10^6$   
determined by using agar surface method (TSA-Oxoid)

Carrier: filter paper  
Packaging: paper / foil

Lot: V 210819

Expiration date: 09.2021

D<sub>121°C</sub>-Value: 2,5 min

z-value: 8,6 °C

Survival time: D-value x (log cfu - 2)

Kill time: D-value x (log cfu + 4)

All data are determined at time of manufacturing.

Resistance data all determined according to ISO 11138 and are only reproducible when exposed and cultured under the exact conditions used to obtain results report here. Results may vary under different test conditions.

Incubation: 7 days at 56 °C ± 2 K in CSB

Storage:

at + 18 °C to + 25 °C and rel. humidity from 35 % to 70 %

Protect from light and sterilants.

Do not use after expiration date. Sterilize all cultures before disposal.

Date: 02.09.2019

Release: T. Seis

SIMICON GMBH • Sigmund-Riefler-Bogen 19 • D - 81829 München  
Tel.: +49 / (0)89 / 67 33 66 0 • Fax: +49 / (0)89 / 67 33 66 22

## SIMICON

Hygiene & Mikrobiologie

SIMICON GMBH  
Sigmund-Riefler-Bogen 19  
81829 München  
Tel.: +49-89-673366-0  
Fax: +49-89-673366-22

02.09.2019

Date

Stamp / Signature