

QUALITY CONTROL CERTIFICATE



Xebios
Diagnostics

Product name : Legionella-BCYE-Agar (with Cystein ISO 11731)
Abbreviation : BCYE+CYS
Article code : 40-1603
Filling Volume : 18 ml
Acceptance date : 23-12-2020
Batch number : 16030518
Storage temp. : 4-25 °C
Expiration date : 15-06-2021
(for sealed enclosure as dd-mm-yyyy)

Typical Formula

*Ingredients in gram per litre of nutrient medium * :*

Yeast extract	10
Activated charcoal	2
ACES buffer	10
L-Cysteine	0,4
Iron (III) pyrophosphate	0,25
Alpha-Ketoglutarate	1
Potassium hydroxide	2,8
Agar	13

** Adjusted as required to meet performance standards.*

Physical Properties

Test	Criterion	Result	Conforms
pH	6,8 ± 0,2	6,8	C
Color	Black	Black	C
Consistency	agar	agar	C
Gel Strength		Conforms	C
Appearance	Clear:No Partic.:No	Clear:No Part.:No	C

Sterility Check

Incubation-condition	Result	Conforms
44-52 hours 30 ± 1 °C	Sterile	C

Growth Properties on next page.

Quality Control Certificate

Product name : Legionella-BCYE-Agar (with Cystein ISO 11731)
Abbreviation : BCYE+CYS Batch number : 16030518
Article code : 40-1603 Storage temp. : 4-25 °C
Acceptance date : 23-12-2020 Expiration date : 15-06-2021

Growth Properties

- Incubation-condition : 116-124 hours 36 ±2 °C Aerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
Legionella pneumophila	ATCC 33152 WDCM 00107 NCTC 11192	04	grey-white	87	82	1.06	≥0.7	C

- * Method 04 = TV 5.04 Non-selective enumeration agar ; Productivity of solid media ; quantitative method (acc. ISO 11133)
Result: Productivity-Ratio (cfu test / cfu ref) ; inoculum 10² cfu
- ** Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium
CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.
C = Conforms X = Does not conform - = Not tested
Keep protected from light and at a constant temperature.
The test results reported on this report were obtained from a sample of the batch submitted to the laboratory.

Xebios is certified according to DIN EN ISO 9001. Our microbiological quality control is carried out by a microbiological laboratory accredited according to EN ISO/IEC 17025. (RVA Registration no.: RvA-L 614). This laboratory certificate is generated electronically and valid without a signature.

Release Date : 28-12-2020
(dd-mm-yyyy)

Crit.version :4 Document ID : Q20201201499