

QUALITY CONTROL CERTIFICATE



Product name : Trypticase Peptone Glucose Yeast Extract Br
Abbreviation : TPGY BROTH
Article code : T490.76.0225
Filling Volume : 225 ml

Batch number : 20200400646
Storage temp. : 2-25 °C
Expiration date : 30-04-2021
(for sealed enclosure as dd-mm-yyyy)

Physical Properties

Test	Criterion	Result	Conforms
pH	7,0 ± 0,2	6,8	C
Color	yellow/brown	yellow/brown	C
Consistency	broth	broth	C
Appearance	Clear:Yes Partic.:No	Clear:Yes Part.:No	C

Sterility Check

Incubation-condition	Result	Conforms
44-52 hours 30 ±1 °C	Sterile	C

Growth Properties

• Incubation-condition : 40-48 hours 37 ±1 °C Anaerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
Clostridium perfringens	ATCC 13124 WDCM 00007 NCTC 6125 /	02	n.a.	n.a.	36	2	≥ 1	C

* Method 02 = TV 5.02 Non-selective enrichment media ; Productivity of liquid media ; qualitative method (acc. ISO 11133)
Result: Turbidity ; 2 = good 1 = slight 0 = no turbidity ; inoculum ≤100 cfu

** Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium
CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.
C = Conforms

Do not expose the product to intense light.

The test results reported on this quality control certificate were obtained from a sample of the batch at the time of release.

Authorisation QC :

Peggy Spaan
11-05-2020
(dd-mm-yyyy)

A handwritten signature in black ink, appearing to be "Peggy Spaan", written over a white background.

Crit.version : 1
Document ID : C20200400646