

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0067 Violet Red Bile Glucose Agar (
Batch Number: 00004223
Expiry Date: 16.12.2023
Date Received: 17.12.2020
Date Tested: 17.12.2020
Sample Condition: Satisfactory
Sample Number: 393523

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP1160 Violet Red Bile Glucose Agar with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Escherichia coli ATCC 8739	96	Dark red/purple colonies with or without precipitation halo	Dark red/purple colonies with or without precipitation halo
Escherichia coli NCTC 12241/ATCC 25922	69	Dark red/purple colonies with or without precipitation halo	Dark red/purple colonies with or without precipitation halo
Salmonella typhimurium NCTC 12023/ATCC 14028	63	Dark red/purple colonies with or without precipitation halo	Dark red/purple colonies with or without precipitation halo

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2 .

Selectivity	Test Result	Specification
Enterococcus faecalis NCTC 12697/ATCC 29212	Inhibited	Inhibited

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0067 Violet Red Bile Glucose Agar (
Batch Number: 00004223
Expiry Date: 16.12.2023
Date Received: 17.12.2020
Date Tested: 17.12.2020
Sample Condition: Satisfactory
Sample Number: 393523

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Physical	Result	Specifications
pH	7.4	7.4 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Light purple98-5 - 146-5

All of the results on this certificate of analysis relate only of the samples submitted.

Douglas Cameron
Technical Manager, E&O Diagnostics Ltd