

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

**Sample:** KM0084 Bacillus cereus Agar (MYP)  
**Batch Number:** 00004620  
**Expiry Date:** 02.09.2024  
**Date Received:** 08.09.2021  
**Date Tested:** 08.09.2021  
**Sample Condition:** Satisfactory  
**Sample Number:** 408866

Product sample prepared and tested as PP0050 MYP with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Bacillus cereus NCTC 10320	113	Pink colonies with white halo	Pink colonies with white halo
Bacillus subtilis NCTC 10400/ATCC 6633	110	Yellow colonies	Yellow colonies

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor $\geq 2$ .		
Selectivity	Test Result	Specification
Escherichia coli NCTC 12241/ATCC 25922	Inhibited	Inhibited

Physical	Result	Specifications
pH	7.0	7.2 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Peach86-4 - 86-7

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** KM0084 Bacillus cereus Agar (MYP)  
**Batch Number:** 00004620  
**Expiry Date:** 02.09.2024  
**Date Received:** 08.09.2021  
**Date Tested:** 08.09.2021  
**Sample Condition:** Satisfactory  
**Sample Number:** 408866

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

---

All of the results on this certificate of analysis relate only of the samples submitted.



**Douglas Cameron**  
Technical Manager, E&O Diagnostics Ltd