

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Batch	Expiration date
TOS Propionate Agar Base	062320503	2024.05.31

Ref. 610378 – 620378

Physical quality control	Specification	Results
Expected pH-value (25°C)	6.7 ± 0.2	6.7
Appearance of powder	Free-flowing, homogeneous	Conforms
Colour of powder	Beige	Conforms
Appearance of prepared medium	Clear	Conforms
Colour of prepared medium	Amber	Conforms

Microbiological Performance

Tested according to CLSI M22-A3, EN ISO 11133
Supplements: MUP Selective Supplement
Reference media: Tryptic Soy Agar

Productivity, Method of control: Quantitative

Inoculum: 50-100 CFU

Control strains	Incubation	Expected Results	Specification	Results
<i>Bifidobacterium animalis</i> subsp. <i>animalis</i> ATCC® 25527	Anaerobic, 72 ± 3 h 37 ± 1°C	$P_R \geq 0.5$	Good growth, white colonies	Conforms
<i>Bifidobacterium animalis</i> subsp. <i>lactis</i> ATCC® 27536	Anaerobic, 72 ± 3 h 37 ± 1°C	$P_R \geq 0.5$	Good growth, white colonies	Conforms

Selectivity, Method of control: Qualitative

Inoculum: 10⁴-10⁶ CFU

Control strains	Incubation	Specification	Results
<i>Lactobacillus casei</i> ATCC® 393	Anaerobic, 72 ± 3 h 37 ± 1°C	Inhibition	Conforms

Batch Release

Approved

Date 30.06.2020

Signature

Quality Control
(D. Vitagliano)

The results reported were obtained at the time of release.