

# QUALITY CONTROL CERTIFICATE



Product name : Legionella-BCYE-Agar (with Cystein ISO 11731)  
Abbreviation : BCYE+CYS  
Article code : 40-1603  
Filling Volume : 18 ml  
Batch number : 16030572  
Storage temp. : 4-25 °C  
Expiration date : 21-10-2021  
(for sealed enclosure as dd-mm-yyyy )

## Typical Formula

Ingredients in gram per litre of nutrient medium \* :

Yeast extract	10
Activated charcoal	2
ACES buffer	10
L-Cysteine	0,4
Iron (III) pyrophosphate	0,25
Alpha-Ketoglutarate	1
Potassium hydroxide	2,8
Agar	13

\* Adjusted as required to meet performance standards.

## Physical Properties

Test	Criterion	Result	Conforms
pH	6,8 ± 0,2	6,7	C
Color	Black	Black	C
Consistency	agar	agar	C
Gel Strength		Conforms	C
Appearance	Clear:No Partic.:No	Clear:No Part.:No	C

## Sterility Check

Incubation-condition	Result	Conforms
44-52 hours 30 ± 1 °C	Sterile	C

## Growth Properties

• Incubation-condition : 116-124 hours 36 ± 2 °C Aerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
Legionella pneumophila	ATCC 33152 WDCM 00107 NCTC 11192	04	grey-white	91	87	1.05	≥ 0.7	C

\* Method 04 = TV 5.04 Non-selective enumeration agar ; Productivity of solid media ; quantitative method (acc. ISO 11133)  
Result: Productivity-Ratio (cfu test / cfu ref) ; inoculum 10<sup>2</sup> cfu

\*\* Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium  
CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.

C = Conforms X = Does not conform - = Not tested

Keep protected from light and at a constant temperature.

The test results reported on this report were obtained from a sample of the batch submitted to the laboratory.

# Quality Control Certificate

Product name : Legionella-BCYE-Agar (with Cystein ISO 11731)  
Abbreviation : BCYE+CYS  
Article code : 40-1603  
Batch number : 16030572  
Storage temp. : 4-25 °C  
Expiration date : 21-10-2021

---

Xebios is certified according to DIN EN ISO 9001. Our microbiological quality control is carried out by a microbiological laboratory accredited according to EN ISO/IEC 17025. (RvA Registration no.: RvA-L 614). This laboratory certificate is generated electronically and valid without a signature. Results were obtained from laboratory report X0210401551.

Release Date :  
03-05-2021  
(*dd-mm-yyyy*)

Crit.version :4 Document ID : Q20210401551