



CERTIFICATE OF CONFORMITY

Certificate for Lot Number: 61705

Product Code: TS/15-B. Product: Blue Sponge in easy open stomacher pouch dosed with 10ml Neutralising Buffer (Pack Size 200)

Date Of Manufacture: 10/2020 Expiry: 2021-10 Shelf Life: 12 months

This is to certify that the product detailed on this certificate conforms to Technical Service Consultants Ltd procedures and specifications and has performed the tests detailed below within documented parameters.

Sterilisation Reference Number

S12490837-1-1 - 9100392 : Dose 25.5 - 48.3 kGy

For Irradiated products batch release is controlled by the verification of the radiation dose throughout the load.

Autoclave Reference Number

For autoclaved products, either a temperature datalogger or visual sterilisation indicator are used to confirm that sterility was achieved.

Physical and Microbiological Assurance

Tests Performed	Results Obtained
Miles/Misra - L.monocytogenes NCTC 12427 (1-20cfu)	Pass
Sponge Inhibition Test - Biocide Free	Pass
Sterility Test from CNTL	Pass
Sterility Release	Pass

Overall Result: Pass

Certificate Issued by: Joanne Walsh Certificate authorised on 19.10.2020

Electronically generated on 26th May 2021 - valid without signature

Sent to Sonja Pallavicini (spa@ssidiagnostica.com)