



CERTIFICATE OF CONFORMITY

Certificate for Lot Number: 61899

Product Code: TS/15-B. Product: Blue Sponge in easy open stomacher pouch dosed with 10ml Neutralising Buffer (Pack Size 200)

Date Of Manufacture: 11/2020 Expiry: 2021-11 Shelf Life: 12 months

This is to certify that the product detailed on this certificate conforms to Technical Service Consultants Ltd procedures and specifications and has performed the tests detailed below within documented parameters.

Sterilisation Reference Number

S12504502-1-1 - 9190026 : Dose 26.5 - 49.5 kGy, S12507079-1-1 - 9204832 : Dose 26.3 - 49 kGy
For Irradiated products batch release is controlled by the verification of the radiation dose throughout the load.

Autoclave Reference Number

For autoclaved products, either a temperature datalogger or visual sterilisation indicator are used to confirm that sterility was achieved.

Physical and Microbiological Assurance

Tests Performed	Results Obtained
Sterility Test from CNTL	Pass
Sterility Release	Pass
Miles/Misra - L.monocytogenes NCTC 12427 (1-20cfu)	Pass
Sponge Inhibition Test - Biocide Free	Pass

Overall Result: Pass

Certificate Issued by: Mia Openshaw Certificate authorised on 23.11.2020

Electronically generated on 26th May 2021 - valid without signature

Sent to Sonja Pallavicini (spa@ssidiagnostica.com)