QUALITY CONTROL CERTIFICATE



Product name : Acetamide Broth

Abbreviation : ACETAMIDE BROTH Batch number : 20211000629

Article code : A501.25.0005 Storage temp. : 2-25 °C Filling Volume : 5 ml Expiration date : 26-10-2022

(for sealed enclosure as dd-mm-yyyy)

Typical Formula

Ingredients per litre of nutrient medium *:

| Potassium dihydrogen phosphate | 1 | gr |
|--------------------------------|------|----|
| Magnesium sulphate | 0.2 | gr |
| Acetamide | 2 | gr |
| Sodium chloride | 0.2 | gr |
| Sodium molybdate dihydrate | 5 | mg |
| Iron (II) sulfate heptahydrate | 0.5 | mg |
| Demiwater | 1000 | ml |

^{*} Adjusted as required to meet peformance standards.

Physical Properties

| Test | Criterium | Result | Conforms |
|--|---|---|-------------|
| pH Color Consistency Appearance | 7.0 ± 0.2 colourless broth Clear: Yes Partic.: No | 7.0 colourless broth Clear:Yes Part.:No | C C C |

Sterility Check

| Incubation-condition | Result | Conforms | |
|---|---------|----------|--|
| 44-52 hours $30 \pm 1 ^{\circ}\text{C}$ | Sterile | C | |

Growth Properties

• Incubation-condition: 20-24 hours 37 ±1 °C Aerobic

| Micro-organisms | Strain | Method* | Electivity** | CFU | CFU-re | f Result | Crit. | Conf. |
|------------------------|------------|---------|--------------|------|--------|----------|-------|-------|
| Pseudomonas aeruginosa | ATCC 27853 | 02 | n.a. | n.a. | 39 | 2 | ≥ 1 | C |
| | WDCM 00025 | | | | | | | |

WDCM 00025 NCTC 12903

- Method 02 = TV 5.02 Non-selective enrichment media ; Productivity of liquid media ; qualitative method (acc. ISO 11133)

 Result: Turbidity ; 2 = good 1 = slight 0 = no turbidity ; inoculum <=100 cfu
- ** Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate. C = Conforms

Do not expose the product to intense light.

The test results reported on this quality control certificate were obtained from a sample of the batch.

Our microbiological quality control is carried out by our microbiological laboratory accredited according to EN ISO/IEC 17025. (RvA Registration number: RvA-L 614). This laboratory certificate is generated electronically and valid without a signature. Results were obtained from laboratory report R20211000629.

Acceptance date: 29-10-2021 Authorisation QC: Michael Rommelse 02-11-2021 (dd-mm-yyyy)

Crit.version:4 Document ID: C20211000629