QUALITY CONTROL CERTIFICATE



Product name : Rappaport Broth (RVS) conform ISO 6579

Abbreviation : RVS Batch number : 20211100532 Article code : R151.25.0010 Storage temp. : 2-25 °C Filling Volume : 10 ml Expiration date : 18-11-2022

(for sealed enclosure as dd-mm-yyyy)

Typical Formula

Ingredients per litre of nutrient medium *:

Papaic digest of soja	4.50	gr
Sodium chloride	7.20	gr
Monopotrassium phosphate	1.26	gr
Dipotassium phosphate	0.18	gr
Magnesium chloride, anhydrous	13.4	gr
Malachite green (oxalate)	36	mg
Demiwater	1000	ml

^{*} Adjusted as required to meet peformance standards.

Physical Properties

Test	Criterium	Result	Conforms
рН	5.2 ± 0.2	5.1	C
Color	blue	blue	C
Consistency	broth	broth	C
Appearance	Clear:Yes Partic.:No	Clear:Yes Part.:No	C

Sterility Check

Incubation-condition	Result	Conforms	
44-52 hours $30 \pm 1 ^{\circ}\text{C}$	Sterile	C	

Growth Properties on next page.

Quality Control Certificate

: Rappaport Broth (RVS) conform ISO 6579 Product name

Abbreviation : RVS Batch number : 20211100532 : 2-25 °C Article code : R151.25.0010 Storage temp.

Expiration date : 18-11-2022

Growth Properties

• Incubation-condition: 21-27 hours 41,5 ±1 °C Aerobic +Transfer: 21-27 hours 37 ±1 °C Aerobic

Micro-organisms 14028+25922+27853	Strain ATCC MIX6	Method*	Electivity** >10 red,black centre XLD	CFU n.a.	CFU-res	f Result	Crit. ≥ 2	Conf.
13076+8739+27853	ATCC MIX5	11	>10 red,black centre XLD	n.a.	15	2	<u>≥</u> 2	C
Enterococcus faecalis	ATCC 29212 WDCM 00087 NCTC 12697	12	<10 colonies	n.a.	22000	0	≤ 1	С
Escherichia coli	ATCC 25922 WDCM 00013 NCTC 12241	12	=<100 colonies	n.a.	22000	0	≤1	С

Measurement uncertainties are available on request.

Method 11 = TV 5.11 Selective enrichment media; Productivity of liquid media; qualitative method (acc. ISO 11133)

Result: Turbidity; 2 = good, 1 = slight, 0 = no turbidity; Target inoculum <=100 cfu, non Target inoculum >=1000 cfu

TV 5.12 Selective enrichment media; Selectivity of liquid media; qualitative method (acc. ISO 11133)

Result: Turbidity ; 2 = good, 1 = slight, 0 = no turbidity ; inoculum 10^3-10^4 cfu

Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.

Do not expose the product to intense light.

The test results reported on this quality control certificate were obtained from a sample of the batch.

CFU Stam ATCC 14028 Salmonella Typhimurium =<100 (=CFU-ref) Target ATCC 25922 Escherichia coli Non-target >1000 ATCC 27853 Pseud. aeruginosa Non-target >1000 MIX 5 Stam CFU ATCC 13076 Salmonella Enteritidis =<100 (=CFU-ref) Target

Non-target >1000 ATCC 8739 Escherichia coli Non-target >1000 ATCC 27853 Pseud. aeruginosa

Our microbiological quality control is carried out by our microbiological laboratory accredited according to EN ISO/IEC 17025. (RvA Registration number: RvA-L 614). This laboratory certificate is generated electronically and valid without a signature. Results were obtained from laboratory report R20211100532.

Crit.version:11 Document ID: C20211100532

Acceptance date: 22-11-2021 Authorisation QC: Leon Thomassen 24-11-2021 (dd-mm-yyyy)