

QUALITY CONTROL CERTIFICATE



Product name : Peptone Physiological Salt Solution
Abbreviation : PFZ
Article code : P100.25.0009
Filling Volume : 9 ml

Batch number : 20211200268
Storage temp. : 2-25 °C
Expiration date : 13-12-2022
(for sealed enclosure as dd-mm-yyyy)

Typical Formula

Ingredients per litre of nutrient medium * :

Tryptone	1	gr
Sodium chloride	8.5	gr
Demiwater	1000	ml

* Adjusted as required to meet performance standards.

Physical Properties

Test	Criterion	Result	Conforms
pH	7.0 ± 0.2	7.0	C
Color	colourless	colourless	C
Consistency	broth	broth	C
Appearance	Clear:Yes Partic.:No	Clear:Yes Part.:No	C
Volume	9,00 ± 2%	9,07 ml 9,06 ml	C

Sterility Check

Incubation-condition	Result	Conforms
44-52 hours 30 ± 1 °C	Sterile	C

Growth Properties

• Incubation-condition : 45 min - 1 hours 25 ± 1 °C Aerobic +Transfer : 40-48 hours 37 ± 1 °C Aerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
Staphylococcus aureus	ATCC 25923 WDCM 00034 NCTC 12981	01	n.a.	9700	8600	1.13	≥ 0.7	C
Escherichia coli	ATCC 25922 WDCM 00013 NCTC 12241	01	n.a.	10000	10800	0.93	≥ 0.7	C

Measurement uncertainties are available on request.

* Method 01 = TV5.01 Non-selective dilution broth ; Productivity of liquid media ; quantitative method (acc. ISO 11133)
Result: Productivity-Ratio (cfu T1 / cfu T0) ; inoculum 10⁴ cfu

** Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium
CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.
C = Conforms

Do not expose the product to intense light.

The test results reported on this quality control certificate were obtained from a sample of the batch.

Our microbiological quality control is carried out by our microbiological laboratory accredited according to EN ISO/IEC 17025. (RvA Registration number: RvA-L 614).

This laboratory certificate is generated electronically and valid without a signature.

Results were obtained from laboratory report R20211200268.

Acceptance date:
15-12-2021

Authorisation QC:
Leon Thomassen
20-12-2021
(dd-mm-yyyy)

Crit.version :13 Document ID : C20211200268