

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: PP0320 - X.L.D. Agar
Batch Number: 04160379
Expiry Date: 2022-06-29
Date Received: 2022-04-21
Date Tested: 2022-04-21
Date of Issue: 2022-05-02
Sample Condition: Satisfactory
Sample Number: 423376

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

The RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media / $\geq 50\%$ for selective media.

Accredited Test Method: ED/SOP/008 Quantitative evaluation using spread inoculum technique (solid media)

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Shigella sonnei NCTC 8574	86	Red colonies	Red colonies
Salmonella typhimurium NCTC 12023	75	Translucent colonies with black centre	Translucent colonies with black centre

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2 .

Selectivity	Test Result	Specification
Escherichia coli NCTC 12241	Inhibited	Inhibited
Enterococcus faecalis NCTC 12697	Inhibited	Inhibited

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: PP0320 - X.L.D. Agar
Batch Number: 04160379
Expiry Date: 2022-06-29
Date Received: 2022-04-21
Date Tested: 2022-04-21
Date of Issue: 2022-05-02
Sample Condition: Satisfactory
Sample Number: 423376

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

Physical	Result	Specifications	Accredited Test Method
Sterility	Conforms	Within acceptable limits	ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C
Sterility sampling is performed in accordance with ISO 2859-1:1999*. The inspection level is $\geq 0.4\%$ of the batch and the reject level ≤ 7 units depending on batch size.			
pH	7.4	7.4 +/- 0.2	ED/SOP/003 measurement by pH meter
Colour	Conforms	60-1 - 78-1 Red	ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide
Fill Quantity	Conforms	18-22ml	ED/SOP/054 by gravimetric determination
Moisture	Conforms	95% +/- 3%	ED/SOP/053 with drying and gravimetric determination

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: PP0320 - X.L.D. Agar
Batch Number: 04160379
Expiry Date: 2022-06-29
Date Received: 2022-04-21
Date Tested: 2022-04-21
Date of Issue: 2022-05-02
Sample Condition: Satisfactory
Sample Number: 423376

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

All of the results on this certificate of analysis relate only to the samples submitted.

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications

*Sterility sampling is outwith scope of accreditation.



1894



Douglas Cameron
Technical Manager, E&O Laboratories Ltd