

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0078 Dichloran Rose Bengal Chloramp
Batch Number: 00004799
Expiry Date: 07.01.2025
Date Received: 14.01.2022
Date Tested: 14.01.2022
Sample Condition: Satisfactory
Sample Number: 416671

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Product sample prepared and tested as PP4253 DRBC with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Saccharomyces cerevisiae NCPF 3178	71	White colonies with pink hue	White colonies with pink hue
Candida albicans NCPF 3179/ATCC 10231	71	White colonies with pink hue	White colonies with pink hue
Aspergillus brasiliensis NCPF 2275/ATCC 16404	80	White colonies with black pigmentation. Reverse side is pale yellow.	White colonies with black pigmentation. Reverse side is pale yellow.
Mucor racemosus DSMZ 5266	82	White colonies with black pigmentation. Reverse side is pale yellow.	White colonies with black pigmentation. Reverse side is pale yellow.

Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor ≥ 2 .		
Selectivity	Test Result	Specification
Escherichia coli NCTC 12241/ATCC 25922	Inhibited	Inhibited
Bacillus subtilis NCTC 10400/ATCC 6633	Inhibited	Inhibited

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



Sample: KM0078 Dichloran Rose Bengal Chloramp
Batch Number: 00004799
Expiry Date: 07.01.2025
Date Received: 14.01.2022
Date Tested: 14.01.2022
Sample Condition: Satisfactory
Sample Number: 416671

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

Physical	Result	Specifications
pH	5.6	5.6 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Deep pink141-1 - 149-1

All of the results on this certificate of analysis relate only of the samples submitted.



Douglas Cameron
Technical Manager, E&O Diagnostics Ltd