

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** KM0067 Violet Red Bile Glucose Agar (  
**Batch Number:** 00004927  
**Expiry Date:** 18.02.2025  
**Date Received:** 07.03.2022  
**Date Tested:** 07.03.2022  
**Sample Condition:** Satisfactory  
**Sample Number:** 420309

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Product sample prepared and tested as PP1160 with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Escherichia coli ATCC 8739	99	Dark red/purple colonies with or without precipitation halo	Dark red/purple colonies with or without precipitation halo
Escherichia coli NCTC 12241/ATCC 25922	86	Dark red/purple colonies with or without precipitation halo	Dark red/purple colonies with or without precipitation halo
Salmonella typhimurium NCTC 12023/ATCC 14028	76	Dark red/purple colonies with or without precipitation halo	Dark red/purple colonies with or without precipitation halo

**Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor  $\geq 2$ .**

Selectivity	Test Result	Specification
Enterococcus faecalis NCTC 12697/ATCC 29212	Inhibited	Inhibited

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** KM0067 Violet Red Bile Glucose Agar (  
**Batch Number:** 00004927  
**Expiry Date:** 18.02.2025  
**Date Received:** 07.03.2022  
**Date Tested:** 07.03.2022  
**Sample Condition:** Satisfactory  
**Sample Number:** 420309

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Physical	Result	Specifications
pH	7.5	7.4 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Light purple98-5 - 146-5

All of the results on this certificate of analysis relate only of the samples submitted.



**Douglas Cameron**  
Technical Manager, E&O Diagnostics Ltd