

QUALITY CONTROL CERTIFICATE



Product name : Nutrient Agar
Abbreviation : NA 1
Article code : N031.77.0400
Filling Volume : 400 ml

Batch number : 20211000487
Storage temp. : 2-25 °C
Expiration date : 20-10-2022
(for sealed enclosure as dd-mm-yyyy)

Typical Formula

Ingredients per litre of nutrient medium * :

Beef extract	3	gr
Peptone	5	gr
Agar	15	gr
Demiwater	1000	ml

* Adjusted as required to meet performance standards.

Physical Properties

Test	Criterion	Result	Conforms
pH	6.8 ± 0.2	6.6	C
Color	yellow	yellow	C
Consistency	agar	agar	C
Gel Strength		Conforms	C
Appearance	Clear:No Partic.:No	Clear:No Part.:No	C

Sterility Check

Incubation-condition	Result	Conforms
44-52 hours 30 ± 1 °C	Sterile	C

Growth Properties on next page.

Quality Control Certificate

Product name : Nutrient Agar
Abbreviation : NA 1
Article code : N031.77.0400

Batch number : 20211000487
Storage temp. : 2-25 °C
Expiration date : 20-10-2022

Growth Properties

• Incubation-condition : 40-48 hours 30 ±1 °C Aerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
Staphylococcus aureus	ATCC 25923 WDCM 00034 NCTC 12981	04	cream-yellow	124	105	1.18	≥ 0.7	C
Bacillus cereus	ATCC 11778 WDCM 00001 NCTC 10320	04	cream	102	97	1.05	≥ 0.7	C
Escherichia coli	ATCC 25922 WDCM 00013 NCTC 12241	04	cream	98	88	1.11	≥ 0.7	C
Pseudomonas aeruginosa	ATCC 27853 WDCM 00025 NCTC 12903	04	green	86	92	0.93	≥ 0.7	C

* Method 04 = TV 5.04 Non-selective enumeration agar ; Productivity of solid media ; quantitative method (acc. ISO 11133)
Result: Productivity-Ratio (cfu test / cfu ref) ; inoculum 10² cfu

** Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium
CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.
C = Conforms

Do not expose the product to intense light.

The test results reported on this quality control certificate were obtained from a sample of the batch.

Our microbiological quality control is carried out by our microbiological laboratory accredited according to EN ISO/IEC 17025. (RvA Registration number: RvA-L 614).
This laboratory certificate is generated electronically and valid without a signature.
Results were obtained from laboratory report R20211000487.

Acceptance date:
22-10-2021
Authorisation QC:
Leon Thomassen
28-10-2021
(dd-mm-yyyy)

Crit.version :3 Document ID : C20211000487