

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** KM0019 Yeast Extract Glucose Chloramp  
**Batch Number:** 00004339  
**Expiry Date:** 23.02.2024  
**Date Received:** 30.03.2021  
**Date Tested:** 30.03.2021  
**Sample Condition:** Satisfactory  
**Sample Number:** 400180

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Product sample prepared and tested as PP1191 Yeast Extract Glucose Chloramphenicol Agar with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
Saccharomyces cerevisiae NCPF 3178	72	White colonies	White colonies
Candida albicans NCPF 3179/ATCC 10231	94	White colonies	White colonies
Aspergillus brasiliensis NCPF 2275/ATCC 16404	102	White colonies with black pigmentation. Reverse side is pale yellow.	White colonies with black pigmentation. Reverse side is pale yellow.

**Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor  $\geq 2$ .**

Selectivity	Test Result	Specification
Escherichia coli NCTC 12241/ATCC 25922	Inhibited	Inhibited

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** KM0019 Yeast Extract Glucose Chloramp  
**Batch Number:** 00004339  
**Expiry Date:** 23.02.2024  
**Date Received:** 30.03.2021  
**Date Tested:** 30.03.2021  
**Sample Condition:** Satisfactory  
**Sample Number:** 400180

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Physical	Result	Specifications
pH	6.6	6.6 +/- 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Pale Straw2-8 - 6-8

All of the results on this certificate of analysis relate only of the samples submitted.



**Douglas Cameron**  
Technical Manager, E&O Diagnostics Ltd