

# QUALITY CONTROL CERTIFICATE



Product name : Legionella-BCYE-Agar (with Cystein ISO 11731) 10 PCS; Petri  
Abbreviation : BCYE+CYS Batch number : 16030800  
Article code : 40-1603 Storage temp. : 4-25 °C  
Filling Volume : 18 ml Expiration date : 20-12-2022  
(for sealed enclosure as dd-mm-yyyy )

## Typical Formula

Ingredients in gram per litre of nutrient medium \* :

Yeast extract	10
Activated charcoal	2
ACES buffer	10
L-Cysteine	0.4
Iron (III) pyrophosphate	0.25
Alpha-Ketoglutarate	1
Potassium hydroxide	2.8
Agar	13

\* Adjusted as required to meet performance standards.

## Physical Properties

Test	Criterion	Result	Conforms
pH	6.8 ± 0.2	6.9	C
Color	Black	Black	C
Consistency	agar	agar	C
Gel Strength		Conforms	C
Appearance	Clear:No Partic.:No	Clear:No Part.:No	C

## Microbial Contamination

Incubation-condition	Result	Conforms
44-52 hours 30 ± 1 °C	No contamination	C

## Growth Properties

• Incubation-condition : 116-124 hours 36 ± 2 °C Aerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
Legionella pneumophila	ATCC 33152 WDCM 00107 NCTC 11192	04	grey-white	93	94	0.99	≥ 0.7	C

\* Method 04 = TV 5.04 Non-selective enumeration agar ; Productivity of solid media ; quantitative method (acc. ISO 11133)  
Result: Productivity-Ratio (cfu test / cfu ref) ; inoculum 10<sup>2</sup> cfu

\*\* Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium  
CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.  
n.a = Not Applicable. Crit.= Criterion C = Conforms  
Keep protected from light and at a constant temperature.  
The test results reported on this quality control certificate were obtained from a sample of the batch.

# Quality Control Certificate

Product name : Legionella-BCYE-Agar (with Cystein ISO 11731) 10 PCS; Petri  
Abbreviation : BCYE+CYS Batch number : 16030800  
Article code : 40-1603 Storage temp. : 4-25 °C  
Expiration date : 20-12-2022

---

Xebios is certified according to DIN EN ISO 9001. Our microbiological quality control is carried out by a microbiological laboratory accredited according to EN ISO/IEC 17025. (RvA Registration no.: RvA-L 614). This laboratory certificate is generated electronically and valid without a signature. Results were obtained from laboratory report X20220604763.

Release Date :  
27-06-2022  
(*dd-mm-yyyy*)

Crit.version :4 Document ID : Q20220604763