

<b>Liofilchem S.r.l.</b>	CERTIFICATO CONTROLLO QUALITÀ <i>QUALITY CONTROL CERTIFICATE</i>	N° 52532 – 525322 - 53532 Revisione 8 del 09.10.2013 Pag. 1 di 1
------------------------------	---	--

PRODOTTO / *PRODUCT* : **CONTACT SLIDE 9**

LOTTO / *BATCH* : **102522070**

DATA PRODUZIONE / *PRODUCTION DATE*: **25.10.2022**

DATA SCADENZA / *EXP. DATE*: **2023.07.22**

**TEST DI STERILITÀ / *STERILITY TEST***

according to ISO 2859-1  
Special Inspection Level :S-3

20 ± 2°C (after 7 days)	-
37°C(after 24 h)	-

**CARATTERISTICHE BIOLOGICHE / *BIOLOGICAL CHARACTERISTICS***

Microbiological growth acc. to ISO 11133-1 and ISO 11133-2 and to CLSI M22-A3

Tempo di incubazione / <i>Incubation time</i>	18 h - 24 h
Temperatura / <i>Temperature</i>	37 °C ± 2 °C
Modalità di incubazione / <i>Procedure of incubation</i>	√ O <sub>2</sub> CO <sub>2</sub>

MEDIUM	PLATE COUNT AGAR + TTC + NEUTRALIZZANTE		PLATE COUNT AGAR + TTC + NEUTRALIZZANTE	
	<i>Growth/Colour</i>		<i>Growth/Colour</i>	
<i>Microorganism</i>	Theoretical results	Result of conformity	Theoretical results	Result of conformity
<i>Escherichia coli</i> ..... ATCC 25922	+ red	<input checked="" type="checkbox"/>	+ red	<input checked="" type="checkbox"/>
<i>Pseudomonas aeruginosa</i> ATCC 27853	+ red	<input checked="" type="checkbox"/>	+ red	<input checked="" type="checkbox"/>
<i>Streptococcus faecalis</i> .....ATCC 19433	+ red	<input checked="" type="checkbox"/>	+ red	<input checked="" type="checkbox"/>
<i>Proteus mirabilis</i> ..... ATCC 25933	+ red	<input checked="" type="checkbox"/>	+ red	<input checked="" type="checkbox"/>
<i>Staphylococcus aureus</i> .... ATCC 25923	+ red	<input checked="" type="checkbox"/>	+ red	<input checked="" type="checkbox"/>
<i>Klebsiella pneumoniae</i> .... ATCC 13883	+ red	<input checked="" type="checkbox"/>	+ red	<input checked="" type="checkbox"/>
pH at 25°C	7.0 ± 0.2	<b>7.0</b>	7.0 ± 0.2	<b>7.0</b>

(+) = *Crescita / Growth*    (-) = *Inibizione / Inhibition*    (±) = *Inibizione parziale/Partial inhibition*

LOTTO / *BATCH* :  Idoneo / Approved

Non Idoneo / Not approved

DATA / *DATE* : 03.11.2022

Quality Control  
(This document has been  
established electronically and is  
valid without signature)