

Liofilchem®

Certificate of Analysis

Page 1 of 1

Product	Ref.	Batch	Expiration date
AD Fosfomycin 0.25-256 µg/mL	77001 – 77061	110822002	2023.03.08

Microbiological Performance
Incubation: Ambient air, 16-20 h, 35 ± 2°C

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Expected Results MIC range (mg/L)	Results MIC (mg/L)
<i>Staphylococcus aureus</i> ATCC® 29213	0.5–4*	2
<i>Enterococcus faecalis</i> ATCC® 29212	32–128	64
<i>Escherichia coli</i> ATCC® 25922	0.5–2*	1
<i>Pseudomonas aeruginosa</i> ATCC® 27853	2–8*	8

*Validated by EUCAST

Batch Release

Approved

Date

16.11.2022

Signature

Quality Control

(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.