

# QUALITY CONTROL CERTIFICATE



Product name : Pseudomonas-CN-Agar  
Abbreviation : PCN  
Article code : 40-1286  
Filling Volume : 18 ml

Batch number : 12860620  
Storage temp. : 4-10 °C  
Expiration date : 13-01-2023  
(for sealed enclosure as dd-mm-yyyy )

## Typical Formula

Ingredients in gram per litre of nutrient medium \* :

Gelatine peptone	16
Casein hydrolysate	10
Potassium sulphate, anhydrous	10
Magnesium chloride, anhydrous	1.4
Agar	11
Cetrimide	0.2
Nalidixic acid [mg]	15
Glycerine [ml]	10

\* Adjusted as required to meet performance standards.

## Physical Properties

Test	Criterion	Result	Conforms
pH	7.1 ± 0.2	7.2	C
Color	light beige	light beige	C
Consistency	agar	agar	C
Gel Strength		Conforms	C
Appearance	Clear:Yes Partic.:No	Clear:Yes Part.:No	C

## Microbial Contamination

Incubation-condition	Result	Conforms
44-52 hours 30 ±1 °C	No contamination	C

**Growth Properties on next page.**

# Quality Control Certificate

Product name : Pseudomonas-CN-Agar  
Abbreviation : PCN  
Article code : 40-1286

Batch number : 12860620  
Storage temp. : 4-10 °C  
Expiration date : 13-01-2023

## Growth Properties

### • Incubation-condition : 40-48 hours 37 ±1 °C Aerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
Pseudomonas aeruginosa	ATCC 10145	06	green, UV fluorescent	108	95	1.14	≥ 0.5	C
	WDCM 00024							
	NCTC 10332							
Enterococcus faecalis	ATCC 19433	07	n.a.	n.a.	16000	0	≤ 0	C
	WDCM 00009							
	NCTC 775							
Escherichia coli	ATCC 8739	07	n.a.	n.a.	23000	0	≤ 0	C
	WDCM 00012							
	NCTC 12923							

- \* Method 06 = TV 5.06 Selective enumeration agar ; Productivity of solid media ; quantitative method (acc. ISO 11133)  
Result: Productivity-Ratio (cfu test / cfu ref) ; inoculum 10<sup>2</sup> cfu
- \* Method 07 = TV 5.07 Selective enumeration agar ; Selectivity of solid media ; qualitative method (acc. ISO 11133)  
Result: Growth ; 2 = good, 1 = slight, 0 = no growth ; inoculum 10<sup>4</sup>-10<sup>6</sup> cfu
- \*\* Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium  
CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.  
n.a = Not Applicable. Crit.= Criterium C = Conforms  
Keep protected from light and at a constant temperature.  
The test results reported on this quality control certificate were obtained from a sample of the batch.

Xebios Diagnostics Group is certified according to DIN EN ISO 9001. Our microbiological quality control is carried out by our EN ISO/IEC 17025 accredited laboratory (RVA Registration no: RvA-L 614) at Tritium Microbiologie BV, Eindhoven.  
Tritium Microbiologie BV is part of Xebios Diagnostics Group.  
This laboratory certificate is generated electronically and valid without a signature.  
Results were obtained from laboratory report X20221000539.

Release Date :  
13-10-2022  
(dd-mm-yyyy)

Crit.version :5 Document ID : Q20221000539