

EU Declaration of Conformity

Manufacturer:	SSI Diagnostica A/S	
	Herredsvejen 2	
	DK-3400 Hillerød	
	Denmark	
SRN no.	DK-MF-000023922	

PRODUCT IDENTIFICATION				
Product Name	Code / Catalog Number			
Dermatophyte and C. albicans	99462			
Real Time PCR kit				
Intended Purpose	Basic UDI-DI			
The Dermatophyte and <i>C. albicans</i> Real Time PCR Kit is an <i>in vitro</i> diagnostic kit for the quantitative detection of general dermatophytes (pan-dermatophytes), <i>Trichophyton (T.) rubrum</i> and <i>Candida (C.) albicans</i> in nail, skin, and hair specimens.	5713106Derm_RT_PCRVV			

RISK CLASS	FOR DEV	ICES
Device Clas	sification	Rule:
Class:	В	IVDR rule 6
	0,00	no / Chandarda

Common Specifications / Standards

SSI Diagnostica A/S declares that the above-mentioned products meet the provision of the following EU legislation:

- European In Vitro Diagnostic Regulation (2017/746)
- ISO 13485:2019 Medical devices Quality management systems -Requirements for regulatory purposes
- ISO 14971:2019 Risk Management
- ISO 13612:2002 Performance evaluation of in vitro diagnostic medical devices
- ISO 15223-1: 2021 Symbols to be used with medical device labels, labelling and information to be supplied Part 1: General requirements
- ISO 18113-1: 2011 Information supplied by the manufacturer (labelling) Part 1: Terms, definitions and general requirements
- ISO 18113-2: 2011 Information supplied by the manufacturer (labelling) In vitro diagnostic reagents for professional use

COMPANY REPRESENTATIVE:

Place/date:

Hilleroed, October 21th 2022

Signature:

Ulla W. Berg

Title:

QA/RA Manager

la/ds