

EU Declaration of Conformity

Manufacturer:	SSI Diagnostica A/S	
	Herredsvejen 2	
	DK-3400 Hillerød	
	Denmark	
SRN no.	DK-MF-000023922	

PRODUCT IDENTIFICATION			
Product Name	Code / Catalog Number		
ImmuLex Pneumococcal antisera	Multiple items - eg. 52390,		
Pool, Group, Type and Factor	77322, 77334, 98925		
Intended Purpose	Basic UDI-DI		
ImmuLex™ pneumococcus Pool, Group, Type and Factor antisera products are intended for visual qualitative confirmation, serogrouping and serotyping of Streptococcus pneumoniae (pneumococcus) by use of a rapid agglutination test. The products are for testing identified and confirmed, pure cultured isolates and strains of pneumococcus	5713106Pneu_ImmuLexDL		

RISK CLASS	FOR DEV	'ICES
Device Clas	sification	Rule:
Class:	В	IVDR rule 6
C		no / Chandauda

Common Specifications / Standards

SSI Diagnostica A/S declares that the above-mentioned products meet the provision of the following EU legislation:

- European In Vitro Diagnostic Regulation (2017/746)
- ISO 13485:2019 Medical devices Quality management systems Requirements for regulatory purposes
- ISO 14971:2019 Risk Management
- ISO 13612:2002 Performance evaluation of in vitro diagnostic medical devices
- ISO 15223-1: 2021 Symbols to be used with medical device labels, labelling and information to be supplied Part 1: General requirements
- ISO 18113-1: 2011 Information supplied by the manufacturer (labelling) Part 1: Terms, definitions and general requirements
- ISO 18113-2: 2011 Information supplied by the manufacturer (labelling) In vitro diagnostic reagents for professional use

COMPANY REPRESENTATIVE:

Place/date: Hillered, October 21th 2022

Signature:

Ulla W. Berg

Title: QA/RA Manager