

EU Declaration of Conformity

Manufacturer:	SSI Diagnostica A/S Herredsvejen 2 DK-3400 Hillerød Denmark
SRN no.	DK-MF-000023922

PRODUCT IDENTIFICATION	
Product Name	Code / Catalog Number
<i>Pneumococcus</i> antisera Omni, Pool, Type, Factor	Multiple items - eg. 2438, 16725, 16922, 16998
Intended Purpose	Basic UDI-DI
The SSI Diagnostica <i>pneumococcus antisera</i> are intended for visual qualitative confirmation and serotyping of <i>Streptococcus pneumoniae</i> (<i>pneumococcus</i>) by use of the Neufeld test (also named the capsular reaction test or the Quellung reaction). This product is for testing of identified and confirmed pure cultured isolates and strains of <i>pneumococcus</i> .	5713106Pneu_SeraBQ

RISK CLASS FOR DEVICES		
Device Classification		Rule:
Class:	B	IVDR rule 6
Common Specifications / Standards		
SSI Diagnostica A/S declares that the above-mentioned products meet the provision of the following EU legislation: <ul style="list-style-type: none"> • European In Vitro Diagnostic Regulation (2017/746) • ISO 13485:2019 - Medical devices - Quality management systems - Requirements for regulatory purposes • ISO 14971:2019 – Risk Management • ISO 13612:2002 – Performance evaluation of in vitro diagnostic medical devices • ISO 15223-1: 2021 - Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements • ISO 18113-1: 2011 - Information supplied by the manufacturer (labelling) – Part 1: Terms, definitions and general requirements • ISO 18113-2: 2011 - Information supplied by the manufacturer (labelling) - In vitro diagnostic reagents for professional use 		

COMPANY REPRESENTATIVE:

Place/date: Hillerød, October 21th 2022

Signature:



Ulla W. Berg

Title:

QA/RA Manager