

EU Declaration of Conformity

| | |
|----------------------|--|
| Manufacturer: | SSI Diagnostica A/S Herredsvejen 2 DK-3400 Hillerød Denmark |
| SRN no. | DK-MF-000023922 |

| PRODUCT IDENTIFICATION | |
|---|---------------------------|
| Product Name | Code / Catalog Number |
| <i>Salmonella</i> Sero-Quick Group kit | 18349 |
| Intended Purpose | Basic UDI-DI |
| The <i>Salmonella</i> Sero-Quick Group Kit is a screening kit for identifying <i>Salmonella</i> isolates to the serogroup level. Sero-Quick Group Kit is used as an in vitro diagnostic aid for qualitative manual complete or partial bacterial serotyping by slide agglutination. It is important to use pure culture isolates for determination of bacterial antigens. | 5713106S_Sero-Q_Grp_KitKL |

| RISK CLASS FOR DEVICES | | |
|---|---|-------------|
| Device Classification | | Rule: |
| Class: | B | IVDR rule 6 |
| Common Specifications / Standards | | |
| SSI Diagnostica A/S declares that the above-mentioned products meet the provision of the following EU legislation: <ul style="list-style-type: none"> European In Vitro Diagnostic Regulation (2017/746) ISO 13485:2019 - Medical devices - Quality management systems - Requirements for regulatory purposes ISO 14971:2019 – Risk Management ISO 13612:2002 – Performance evaluation of in vitro diagnostic medical devices ISO 15223-1: 2021 - Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements ISO 18113-1: 2011 - Information supplied by the manufacturer (labelling) – Part 1: Terms, definitions and general requirements ISO 18113-2: 2011 - Information supplied by the manufacturer (labelling) - In vitro diagnostic reagents for professional use | | |

COMPANY REPRESENTATIVE:

Place/date: Hillerød, October 21th 2022

Signature: 
 Title: Ulla W. Berg
 QA/RA Manager