

QUALITY CONTROL CERTIFICATE



Product name : Galle-Äsculin-Azid-Agar
Abbreviation : GÄA
Article code : 40-1276
Filling Volume : 18 ml

Batch number : 12760164
Storage temp. : 4-10 °C
Expiration date : 29-03-2023
(for sealed enclosure as dd-mm-yyyy)

Physical Properties

Test	Criterion	Result	Conforms
pH	7.1 ± 0.2	7.1	C
Color	yellowish	yellowish	C
Consistency	agar	agar	C
Gel Strength		Conforms	C
Appearance	Clear:Yes Partic.:No	Clear:Yes Part.:No	C

Microbial Contamination

Incubation-condition	Result	Conforms
44-52 hours 30 ± 1 °C	No contamination	C

Growth Properties

• Incubation-condition : 40-48 hours 37 ± 1 °C Aerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
Enterococcus faecalis	ATCC 19433 WDCM 00009 NCTC 775	06	grey-black	116	100	1.16	≥ 0.5	C
Enterococcus faecium	NCTC 13169 WDCM 178	06	grey-black	140	114	1.23	≥ 0.5	C
Escherichia coli	ATCC 8739 WDCM 00012 NCTC 12923	07	n.a.	n.a.	58000	0	≤ 0	C
Staphylococcus aureus	ATCC 25923 WDCM 00034 NCTC 12981	08	white	n.a.	8000	1	≥ 1	C

* Method 06 = TV 5.06 Selective enumeration agar ; Productivity of solid media ; quantitative method (acc. ISO 11133)
Result: Productivity-Ratio (cfu test / cfu ref) ; inoculum 10² cfu

* Method 07 = TV 5.07 Selective enumeration agar ; Selectivity of solid media ; qualitative method (acc. ISO 11133)
Result: Growth ; 2 = good, 1 = slight, 0 = no growth ; inoculum 10⁴-10⁶ cfu

* Method 08 = TV 5.08 Selective enumeration agar ; Specificity of solid media ; qualitative method (acc. ISO 11133)
Result: Growth ; 2 = good, 1 = slight, 0 = no growth ; inoculum 10³-10⁴ cfu

** Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium
CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.
n.a = Not Applicable. Crit.= Criterion C = Conforms

Keep protected from light and at a constant temperature.

The test results reported on this quality control certificate were obtained from a sample of the batch.

Xebios Diagnostics Group is certified according to DIN EN ISO 9001. Our microbiological quality control is carried out by our EN ISO/IEC 17025 accredited laboratory (RVA Registration no: RvA-L 614) at Tritium Microbiologie BV, Eindhoven.
Tritium Microbiologie BV is part of Xebios Diagnostics Group.
This laboratory certificate is generated electronically and valid without a signature.
Results were obtained from laboratory report X20221103661.

Release Date :
02-12-2022
(dd-mm-yyyy)

Crit.version :4 Document ID : Q20221103661