

HiMedia Laboratories Private Limited

C-40, Road No.21Y, MIDC, Wagle Ind. Area, Thane(W) - 400604

Website : www.himedialabs.com, Email : info@himedialabs.com

Certificate of Analysis , Quality and Conformity

Material Code : R096	Material Name : Acid Phosphatase Reagent	Lot No : 0000574437
AR No.: 40001324498	Date of Report : 2023-02-04	Exp. Date : 2023-07

TEST	SPECIFICATIONS	RESULTS
<u>Appearance</u> 1 Appearance	Light yellow to brown colour solution	Complies
<u>Clarity</u> 1 Clarity	Hazy solution with precipitate. Note : Precipitate will not affect the performance criteria of the reagent.	Complies
<u>Cultural Response</u> 1 Cultural response	Add 0.2 - 0.3 ml of Acid Phosphatase reagent to a 24 - 48 hours old culture of the organism under investigation.	-
<u>Clostridium perfringens ATCC 12924</u> 1 Acid Phosphatase	strong purple colour developed within 3-4min	Complies
<u>Clostridium perfringens ATCC 13124 (WDCM 00007)</u> 1 Acid Phosphatase	strong purple colour developed within 3-4 min	Complies
<u>Escherichia coli ATCC 25922 (WDCM 00013)</u> 1 Acid Phosphatase	No colour change	Complies

STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current HiMedia manual or product sheets.

This document has been produced electronically and is valid

HiMedia Laboratories Private Limited

C-40, Road No.21Y, MIDC, Wagle Ind. Area, Thane(W) - 400604

Website : www.himedialabs.com, Email : info@himedialabs.com

Material Code : R096	Material Name : Acid Phosphatase Reagent	Lot No : 0000574437
AR No.: 40001324498	Date of Report : 2023-02-04	Exp. Date : 2023-07

TEST	SPECIFICATIONS	RESULTS
 Sukanya Rahate Microbiologist/Sr.Executive Microbiologist	 Ujwala M. Kokate Asst./Dy/QC Manager	 Dr. Santosh Kaul Dy/QA Manager