

EU Declaration of Conformity

| | |
|----------------------|--|
| Manufacturer: | SSI Diagnostica A/S Herredsvejen 2 DK-3400 Hillerød Denmark |
| SRN no. | DK-MF-000023922 |

| PRODUCT IDENTIFICATION | |
|---|--------------------------|
| Product Name | Code / Catalog Number |
| ImmuView <i>S. pneumoniae</i> and <i>L. pneumophila</i> Urinary Antigen Test | 2451 & 10557 |
| Intended Purpose | Basic UDI-DI |
| <p>The ImmuView® <i>S. pneumoniae</i> and <i>L. pneumophila</i> Urinary Antigen Test is an in vitro rapid lateral flow test, also known as a lateral flow immunochromatographic assay.</p> <p>The assay is intended for qualitative detection of <i>Streptococcus (S.) pneumoniae</i> and <i>Legionella (L.) pneumophila</i> antigens in urine specimens from patients with symptoms of pneumonia.</p> <p>The ImmuView® <i>S. pneumoniae</i> and <i>L. pneumophila</i> Urinary Antigen Test can be read visually or used in conjunction with the ImmuView® reader.</p> <p>The assay is intended to aid in diagnosis of <i>S. pneumoniae</i> and of <i>L. pneumophila</i> serogroup 1 infections.</p> <p>The assay is further intended to aid in the diagnosis of <i>S. pneumoniae</i> infections by detection of <i>S. pneumoniae</i> antigen in cerebrospinal fluid (CSF).</p> <p>Results from the ImmuView® <i>S. pneumoniae</i> and <i>L. pneumophila</i> Urinary Antigen Test should be interpreted in conjunction with the patient's clinical evaluation and other diagnostic methods.</p> | 5713106ImmuView_PutLutZH |

| RISK CLASS FOR DEVICES | | |
|--|----------|-----------------|
| Device Classification | | Rule: |
| Class: | C | IVDR rule 3 (b) |
| Common Specifications / Standards | | |
| <p>SSI Diagnostica A/S declares that the above-mentioned products meet the provision of the following EU legislation:</p> <ul style="list-style-type: none">• European In Vitro Diagnostic Regulation (2017/746)• ISO 13485:2019 - Medical devices - Quality management systems - Requirements for regulatory purposes• ISO 14971:2019 – Risk Management• ISO 13612:2002 – Performance evaluation of in vitro diagnostic medical devices• ISO 15223-1: 2021 - Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements• ISO 18113-1: 2011 - Information supplied by the manufacturer (labelling) – Part 1: Terms, definitions and general requirements• ISO 18113-2: 2011 - Information supplied by the manufacturer (labelling) - In vitro diagnostic reagents for professional use | | |

COMPANY REPRESENTATIVE:

Place/date: Hilleroed, February 21th 2023

Signature: 
Ulla W. Berg

Title: QA/RA Manager