

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** KM0019 Yeast Extract Glucose Chloramphenicol Agar  
**Batch Number:** 00005567  
**Expiry Date:** 06.01.2026  
**Date Received:** 11.01.2023  
**Date Tested:** 11.01.2023  
**Sample Condition:** Satisfactory  
**Sample Number:** 439070

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Product sample prepared and tested as PP1191 Yeast Extract Glucose Chloramphenicol with the appropriate selective and/or enrichment supplements added.

Productivity	RGI (%)	Colonial Appearance	Colonial Appearance Specification
C.albicans NCPF 3179	87	White colonies	White colonies
A.brasiliensis NCPF 2275	95	White colonies with black pigmentation, reverse side pale yellow	White colonies with black pigmentation, reverse side pale yellow
S.cerevisiae NCPF 3178	112	White colonies	White colonies

**Accredited Test Method: ED/SOP/051 - log reduction of inoculum using drop inoculum for selective agar and fluid media. Inhibition indicates selectivity factor  $\geq 2$ .**

Selectivity	Test Result	Specification
XB-E.coli NCTC 12241	Inhibited	Inhibited

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** KM0019 Yeast Extract Glucose Chloramphenicol Agar  
**Batch Number:** 00005567  
**Expiry Date:** 06.01.2026  
**Date Received:** 11.01.2023  
**Date Tested:** 11.01.2023  
**Sample Condition:** Satisfactory  
**Sample Number:** 439070

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Physical	Result	Specifications
pH	6.4	6.6 ± 0.2
Appearance	Conforms	Free flowing powder
Colour of Powder	Conforms	Straw
Colour on reconstitution	Conforms	Pale straw. 2-8 – 6-8

All of the results on this certificate of analysis relate only of the samples submitted.



**Douglas Cameron**  
Technical Manager, E&O Diagnostics Ltd