Certificate of Analysis

Client

E&O Laboratories Ltd Burnhouse Bonnybridge Scotland FK4 2HH

$\overline{\ }$	
E&O Labora	atories Ltd

	A source distant Tarata Adaption and Constantial in a large	
Sample Number:	442272	
Sample Condition:	Satisfactory	
Date of Issue:	2023-03-03	
Date Tested:	2023-02-28	
Date Received:	2023-02-28	
Expiry Date:	2024-02-27	Email: info@eolabs.com
Batch Number:	04103367	Telephone: 01324 840404 Fax:01324 841314
Sample:	BM0082 - Buffered Peptone Water(ISO)	Burnhouse, Bonnybridge Scotland, FK4 2HH

Accredited Test Method: See table below			
Productivity	Result	Colonial Appearance	Colonial Appearance Specification
Salmonella typhimurium NCTC 12023	Growth	Yellow colonies	Yellow colonies
Staphylococcus aureus NCTC 12981	Growth	Pink colonies	Pink colonies
Escherichia coli NCTC 12241	Growth	Red colonies	Red colonies

For solid media the RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media / >= 50% for selective media. For liquid media growth is determined from a subculture of the test broth inoculated with 10 - 100 cfu of the target organism.

The log difference between	inocula pre and post a 45 minute retentio Accredited Test Method: See table belo	
Productivity	Result	Specification
Escherichia coli NCTC 12241	Conforms	± 30%
Listeria monocytogenes NCTC 11994	Conforms	± 30%
Staphylococcus aureus NCTC 12981	Conforms	± 30%

Certificate of Analysis

<u>Client</u>

E&O Laboratories Ltd Burnhouse Bonnybridge Scotland FK4 2HH

Sample: Batch Number: Expiry Date: Date Received: Date Tested: Date of Issue: Sample Condition Sample Number:	04103367 2024-02-27 2023-02-28 2023-02-28 2023-03-03 n: Satisfactory		Burnhouse, Bonnybridge Scotland, FK4 2HH Telephone: 01324 840404 Fax:01324 841314 Email: info@eolabs.com
Physical	Result	Specifications	Accredited Test Method
Sterility	Conforms	Within acceptable limits	ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C
Sterility sampling is performed in accordance with ISO 2859-1:1999*. The inspection level is ≥ 0.4% of the batch and the reject level ≤ 7 units depending on batch size.			
рН	7.1	7.0 +/- 0.2	ED/SOP/003 measurement by pH meter
Colour	Conforms	600 C - 4017 C Straw	ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide
Fill Quantity			ED/SOP/054 by gravimetric determination



Page 2 of 3

Certificate of Analysis

<u>Client</u>

E&O Laboratories Ltd Burnhouse Bonnybridge Scotland FK4 2HH

Sample:	BM0082 - Buffered Peptone Water(ISO)
Batch Number:	04103367
Expiry Date:	2024-02-27
Date Received:	2023-02-28
Date Tested:	2023-02-28
Date of Issue:	2023-03-03
Sample Condition:	Satisfactory
Sample Number:	442272



Burnhouse, Bonnybridge Scotland, FK4 2HH Telephone: 01324 840404 Fax:01324 841314 Email: info@eolabs.com

Accredited Test Method for Solid Media

ED/SOP/008 Quantitative evaluation using spread inoculum technique

Accredited Test Method for Liquid Media ED/SOP/008 Semi quantitative inoculation with growth assessment from subculture Accredited Test Method for Diluents

ED/SOP/008 Semi quantitative evaluation of viability maintenance after a holding time of 45 minutes

All of the results on this certificate of analysis relate only to the samples submitted.

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications

*Sterilitv sampling is outwith scope of accreditation.



m

Douglas Cameron Technical Manager, E&O Laboratories Ltd