

# **CERTIFICATE OF ANALYSIS**

I. **Product Information:** 

Product name:

FOB Rapid Test (Cassette) (50 ng/ml)

Catalog No.:

R2010C

Lot No. :

F0301U2D00D 2023.03.04

Manufacture date:

**Expiration date:** 

2025.03.03

**Extraction buffer:** 

Catalog: SB-R2010, Lot: S014722-161, Exp.: 2026.01.04

Storage:

Room temperature (2-30°C)

### Applicable Inspection and Evaluation of Specification: II.

## 1. Physical Inspection:

| Te | st Description         | Acceptance Criteria | <u>Test Results</u> |  |
|----|------------------------|---------------------|---------------------|--|
| a. | Pouch                  | Ac ≤ 5              | 0                   |  |
| b. | Label                  | Ac≤5                | 0                   |  |
| C. | Insert                 | Ac ≤ 5              | 0                   |  |
| d. | Extraction buffer      | Ac ≤ 5              | 0                   |  |
| e. | Patient ID stickers    | Ac ≤ 5              | 0                   |  |
| f. | Desiccant in the Pouch | Ac ≤ 5              | 0                   |  |
| g. | Cassette in the Pouch  | Ac ≤ 5              | 0                   |  |
| h. | Label on the Cassette  | Ac ≤ 5              | 0                   |  |
|    |                        |                     |                     |  |

AC (Acceptance Number) is the maximum allowable defectives in a sample for the lot to be accepted based on the sampling plan.

# 2. Performance Characteristic Inspection:

| Test Description |   | Acceptance  |   | lest Results                                       |  |
|------------------|---|---|---|--|--|
| _                |   | Level   | Criteria  |  |  |
| a.<br>b.         | Negativity Positivity                           | 10 negative samples 10 positive samples                                       | All negative All positive                                   | All negative All positive                          |  |
| c.<br>d.         | Precision<br>LOD                                | 10 replicates<br>R2010-L1(100ng/ml)<br>R2010-L2(50ng/ml)<br>R2010-L3(25ng/ml) | Uniformed band<br>Positive<br>Positive<br>Positive/Negative | Uniformed band<br>Positive<br>Positive<br>Positive |  |
| e.<br>f.<br>g.   | Hook effect(4mg/ml) Extraction buffer Flow Rate | 3 replicates<br>SB-R2010<br>Completion of run in membrane                     | All positive<br>Negative<br>In 2 minutes                    | All positive<br>Negative<br>In 2 minutes           |  |

#### 111. Comments:

All data are from Q.C. test unless otherwise indicated.

### IV. Safety Certification:

There is no biohazard materials utilized in this product, no radioactive materials, no narcotic reagents, neither their derivatives, no virus neither their reagent, nor metabolic by-products.

**Disposition:** V.

'Released [ ✓ ]

Rejected [ ]

Sandy Liu

2023.03.14

Sandy Liu(QA representative)