

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** BM0760 - Max. Recovery Diluent  
**Batch Number:** 04104442  
**Expiry Date:** 2023-11-23  
**Date Received:** 2023-05-26  
**Date Tested:** 2023-05-26  
**Date of Issue:** 2023-05-29  
**Sample Condition:** Satisfactory  
**Sample Number:** 447584

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax: 01324 841314  
Email: info@eolabs.com

**The log difference between inocula pre and post a 45 minute retention in the diluent should be  $\pm 30\%$ .**

**Accredited Test Method: See table below**

Productivity	Result	Specification
Escherichia coli NCTC 12241	Conforms	$\pm 30\%$
Staphylococcus aureus NCTC 12981	Conforms	$\pm 30\%$

Physical	Result	Specifications	Accredited Test Method
Sterility	Conforms	Within acceptable limits	ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C
Sterility sampling is performed in accordance with ISO 2859-1:1999*. The inspection level is $\geq 0.4\%$ of the batch and the reject level $\leq 7$ units depending on batch size.			
pH	7.1	7.0 +/- 0.2	ED/SOP/003 measurement by pH meter
Colour	Conforms	Colourless	ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide
Fill Quantity	Conforms	9ml +/- 0.3	ED/SOP/054 by gravimetric determination

# Certificate of Analysis

## Client

E&O Laboratories Ltd  
Burnhouse  
Bonnybridge  
Scotland  
FK4 2HH



**Sample:** BM0760 - Max. Recovery Diluent  
**Batch Number:** 04104442  
**Expiry Date:** 2023-11-23  
**Date Received:** 2023-05-26  
**Date Tested:** 2023-05-26  
**Date of Issue:** 2023-05-29  
**Sample Condition:** Satisfactory  
**Sample Number:** 447584

Burnhouse, Bonnybridge  
Scotland, FK4 2HH  
Telephone: 01324 840404  
Fax:01324 841314  
Email: info@eolabs.com

Accredited Test Method for Solid Media
ED/SOP/008 Quantitative evaluation using spread inoculum technique

Accredited Test Method for Liquid Media
ED/SOP/008 Semi quantitative inoculation with growth assessment from subculture

Accredited Test Method for Diluents
ED/SOP/008 Semi quantitative evaluation of viability maintenance after a holding time of 45 minutes

All of the results on this certificate of analysis relate only to the samples submitted.

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications

\*Sterility sampling is outwith scope of accreditation.



1894



Douglas Cameron  
Technical Manager, E&O Laboratories Ltd