

**HiMedia Laboratories Private Limited**

C-40, Road No.21Y, MIDC, Wagle Industrial Area,  
Thane(W) - 400604 , Website : www.himedialabs.com,  
Email : info@himedialabs.com

**Certificate of Analysis, Quality and Conformity**

|                                |   |                              |
|--------------------------------|---|------------------------------|
| <b>Material Code</b> : DD018   | <b>Material Name</b> :<br>Oxidase Discs (50 discs / vial) | <b>Lot No</b> : 000584147    |
| <b>Report No.:</b> 40001344850 | <b>Date of Release &amp; Report</b> : 2023-04-12          | <b>Expiry Date</b> : 2024-04 |

**Appearance**

Filter paper discs of 10 mm diameter

**Cultural response**

Typical oxidase reaction given by 18-48 hours old grown culture & observed within 5-10 seconds at 25-30°C.

| Organism                                 | Reaction Observed  |
|--|--|
| <b>Cultural response</b>                 |  |
| <i>Pseudomonas aeruginosa</i> ATCC 27853 | positive : deep purplish blue colouration of disc                    |
| <i>Neisseria gonorrhoeae</i> ATCC 19424  | positive : deep purplish blue colouration of disc                    |
| <i>Escherichia coli</i> ATCC 25922       | negative : purplish blue colouration after 10 sec / no colour change |
| <i>Staphylococcus aureus</i> ATCC 25923  | negative: purplish blue colouration after 10 sec / no colour change  |

- . ATCC is a registered trade mark of the American Type Culture Collection
- . NCTC and National Collection of Type Culture are registered trade mark of the Health Protection Agency

**Control Media :**

- . For Bacteria : Soyabean Casein Digest Agar / Columbia Blood Agar base enriched with 5% v/v Sheep/Horse blood.
- . For Yeast & Mold : Sabouraud Dextrose Agar.
- . All ISO 11133 : 2014/Amd.1:2018( E ) control strains are included in the Quality parameter
- . HiMedia Laboratories Pvt Ltd is Certified for ISO 9001:2015, ISO 13485:2016 , WHO GMP

**Storage & Shelf Life**

Store between 2 - 8°C. Use before expiry date on the label.

**STATUS OF THE MATERIAL : APPROVED**

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets. The results reported were obtained at the time of release.

**This document has been produced electronically and is valid**

  
Pooja Janak Joshi

**Microbiologist/Sr.Executive  
Microbiologist**

  
Ujwala M. Kokate

**Asst./Dy/QC Manager**

  
Dr. Santosh Kaul

**Dy/QA Manager**

--