



Certificate of Analysis: Lyophilized Microorganism Specification and Performance Upon Release

<b>Specifications</b> <b>Microorganism Name:</b> Candida albicans <b>Catalog Number:</b> 0443 <b>Lot Number:</b> 443-1497** <b>Reference Number:</b> ATCC® 10231™* <b>Passage from Reference:</b> 3 <b>(7) Mean Assay Value (MAV):</b> 6.1E+02 CFU per pellet	<b>Expiration Date:</b> 2025/2/28 <b>Release Information:</b> <b>Quality Control Technologist:</b> Mariah H Smith <b>Release Date:</b> 2023/3/27
---	---

<b>Performance</b>	
<b>Macroscopic Features:</b> Small to medium, white, circular, convex, dull colonies. <b>Microscopic Features:</b> Gram positive, ovoidal, budding yeast cells.	<b>Medium:</b> Nutrient <b>Method:</b> Gram Stain (1)

**ID System:** MALDI-TOF (1)  
 See attached ID System results document.

  
 Amanda Kuperus  
 Director of Quality Control  
 AUTHORIZED SIGNATURE

\*\*Disclaimer: The last digit(s) of the lot number appearing on the product label and packing slip are merely a packaging event number. The lot number displayed on this certificate is the actual base lot number.

⚠ Refer to the enclosed product insert for instructions, intended use and hazard/safety information.

Individual products are traceable to a recognized culture collection.



(\*) The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative word mark and the ATCC catalog marks are trademarks of ATCC. Microbiologics, Inc. is licensed to use these trademarks and to sell products derived from ATCC® cultures.

(1) These tests are accredited to ISO/IEC 17025.



(7) The Mean Assay Value (MAV) stated above may deviate from the end-user's MAV based on variables inherent to each laboratory environment, such as methods, media type, equipment, pipettes, and individual technician technique.

# Bruker Daltonik MALDI Biotyper Classification Results



## Meaning of Score Values

Range	Interpretation	Symbols	Color
2.00 – 3.00	High-confidence identification	(+++)	green
1.70 – 1.99	Low-confidence identification	(+)	yellow
0.00 – 1.69	No Organism Identification Possible	(-)	red

## Meaning of Consistency Categories (A - C)

Category	Interpretation
(A)	<b>High consistency:</b> The best match is a high-confidence identification. The second-best match is (1) a high-confidence identification in which the species is identical to the best match, (2) a low-confidence identification in which the species or genus is identical to the best match, or (3) a non-identification.
(B)	<b>Low consistency:</b> The requirements for high consistency are not met. The best match is a high- or low-confidence identification. The second-best match is (1) a high- or low-confidence identification in which genus is identical to the best match or (2) a non-identification.
(C)	<b>No consistency:</b> The requirements for high or low consistency are not met.

Run Creation Date/Time: 2023-03-27T11:37:43.796 MHS

Applied MSP Library(ies): BDAL, Mycobacteria Library (bead method), Filamentous Fungi Library

Sample Name	Sample ID	Organism (best match)	Score Value
B5 (+++) (A)	443-1497	Candida albicans	2.37

Comments:

N/A
-----



## Statistical Analysis Certificate

Microorganism Name: *Candida albicans*

Reference #: ATCC® 10231™\*

Catalog #: 0443

Lot #: 443-1497\*\*

Expiration Date: 2025/2/28

(7) Mean Assay Value (MAV): 6.1E+02 CFU per pellet

Standard Deviation: 1.8E+02

Coefficient of Variation: 30 %

99% Confidence Interval of 5.4E+02 to 6.8E+02 CFU

95% Confidence Interval of 5.6E+02 to 6.7E+02 CFU

Method used to determine Mean Assay Value: Spiral Plate Method

Medium Employed: TSA

Incubation Time and Temp: 48-72 hrs at 28-32 degrees C

A handwritten signature in black ink that reads "Amanda Kuperus". The signature is fluid and cursive.

Amanda Kuperus

Director of Quality Control

AUTHORIZED SIGNATURE

(7) The Mean Assay Value (MAV) stated above may deviate from the end-user's MAV based on variables inherent to each laboratory environment, such as methods, media type, equipment, pipettes, and individual technician technique.