

QUALITY CONTROL CERTIFICATE



Product name : Rappaport Broth (RVS) conform ISO 6579
Abbreviation : RVS
Article code : R151.25.0010
Filling Volume : 10 ml

Batch number : 20230800016
Storage temp. : 2-25 °C
Expiration date : 31-07-2024
(for sealed enclosure as dd-mm-yyyy)

Typical Formula

*Ingredients per litre of nutrient medium * :*

Papaic digest of soja	4.50	gr
Sodium chloride	7.20	gr
Monopotassium phosphate	1.26	gr
Dipotassium phosphate	0.18	gr
Magnesium chloride, anhydrous	13.4	gr
Malachite green (oxalate)	36	mg
Demiwater	1000	ml

** Adjusted as required to meet performance standards.*

Physical Properties

Test	Criterion	Result	Conforms
pH	5.2 ± 0.2	5.0	C
Color	blue	blue	C
Consistency	broth	broth	C
Appearance	Clear:Yes Partic.:No	Clear:Yes Part.:No	C

Microbial Contamination

Incubation-condition	Result	Conforms
44-52 hours 30 ±1 °C	No contamination	C

Growth Properties on next page.

Quality Control Certificate

Product name : Rappaport Broth (RVS) conform ISO 6579
Abbreviation : RVS
Article code : R151.25.0010

Batch number : 20230800016
Storage temp. : 2-25 °C
Expiration date : 31-07-2024

Growth Properties

• Incubation-condition : 21-27 hours 41,5 ±1 °C Aerobic +Transfer : 21-27 hours 37 ±1 °C Aerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
14028+25922+27853	ATCC MIX6	11	>10 red,black centre XLD	n.a.	51	2	≥2	C
13076+8739+27853	ATCC MIX5	11	>10 red,black centre XLD	n.a.	64	2	≥2	C
Enterococcus faecalis	ATCC 29212 WDCM 00087 NCTC 12697	12	<10 colonies	n.a.	78000	0	≤1	C
Escherichia coli	ATCC 25922 WDCM 00013 NCTC 12241	12	≤<100 colonies	n.a.	75000	1	≤1	C

Measurement uncertainties are available on request.

- * Method 11 = TV 5.11 Selective enrichment media ; Productivity of liquid media ; qualitative method (acc. ISO 11133)
Result: Turbidity ; 2 = good, 1 = slight, 0 = no turbidity ; Target inoculum ≤100 cfu, non Target inoculum ≥1000 cfu
- * Method 12 = TV 5.12 Selective enrichment media ; Selectivity of liquid media ; qualitative method (acc. ISO 11133)
Result: Turbidity ; 2 = good, 1 = slight, 0 = no turbidity ; inoculum 10³-10⁴ cfu

** Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium
CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.
n.a = Not Applicable. Crit.= Criterium C = Conforms

Do not expose the product to intense light.

The test results reported on this quality control certificate were obtained from a sample of the batch.

MIX 6	Stam	CFU
ATCC 14028 Salmonella Typhimurium	Target	≤<100 (=CFU-ref)
ATCC 25922 Escherichia coli	Non-target	>1000
ATCC 27853 Pseud. aeruginosa	Non-target	>1000
MIX 5	Stam	CFU
ATCC 13076 Salmonella Enteritidis	Target	≤<100 (=CFU-ref)
ATCC 8739 Escherichia coli	Non-target	>1000
ATCC 27853 Pseud. aeruginosa	Non-target	>1000

Xebios Diagnostics Group is certified according to DIN EN ISO 9001. Our microbiological quality control is carried out by our EN ISO/IEC 17025 accredited laboratory (RVA Registration no: RvA-L 614) at Tritium Microbiologie BV, Eindhoven.
Tritium Microbiologie BV is part of Xebios Diagnostics Group.
This laboratory certificate is generated electronically and valid without a signature.
Results were obtained from laboratory report R20230800016.

Release Date :
08-08-2023
(dd-mm-yyyy)

Crit.version :11 Document ID : C20230800016