

HiMedia Laboratories Private Limited

C-40, Road No.21Y, MIDC, Wagle Industrial Area, Thane(W) - 400604, Website: www.himedialabs.com,

Email: info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M1801	Material Name : Gifu Anaerobic Broth (GAM)	Lot No : 0000573598
Report No.: 40001323092	Date of Release & Report : 2023-01-30	Expiry Date : 2027-12

Appearance

Light yellow to brownish yellow homogeneous free flowing powder. Observed: Light yellow

Colour and Clarity of prepared medium

Amber coloured Clear solution (may have slight precipitate) forms in tube

Reaction

Reaction of 5.9% w/v aqueous solution at 25°C.

Hq

pH Range: 7.20-7.40 Observed: 7.38

Cultural Response

Cultural characteristics observed in an anaerobic atmosphere after an in cubation at 35 - 37°C for 48 - 72 hours.

Organism	Inoculum (CFU)	Growth		
Cultural Response				
Streptococcus pyogenes ATCC 19615	50-100	good - luxuriant		
Bacteroides vulgatus ATCC 8482	50-100	good - luxuriant		
Clostridium sporogenes ATCC 11437	50-100	good - luxuriant		
Clostridium perfringens ATCC 13124 (WDCM	50-100	good - luxuriant		
00007)				

- . ATCC is a registered trade mark of the American Type Culture Collection
- . NCTC and National Collection of Type Culture are registered trade mark of the Health Protection Agency
- . All ISO 11133 : 2014/Amd.1:2018(E) control strains are included in the Quality parameter
- . HiMedia Laboratories Pvt Ltd is Certified for ISO 9001:2015, ISO 13485:2016, WHO GMP

. Information for BSE/TSE Risk The material was subjected to pH <= 7.0 and/or a temperature in excess of 75°C for no less than 2 hours during the manufacturing process. The bovine raw material for this product was collected entirely from Indian Origin animals in a licensed based establishment. The animals are inspected under a Govt. approved veterinarian's supervision and were apparently free from infectious and contagious diseases. BSE (Bovine Spongiform Encephalopathy)/ TSE (Transmissible Spongiform Encephalopathy) and dioxine are not known to exist in India. This material does not contain, nor is derived from the specific risks material as defined in The Maharashtra Animal Preservation Act Govt. of Maharashtra, India.

STATUS OF THE MATERIAL: APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications. The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use,

PAGE: 1/2





HiMedia Laboratories Private Limited

C-40, Road No.21Y, MIDC, Wagle Industrial Area, Thane(W) - 400604, Website: www.himedialabs.com,

Email: in fo@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M1801	Material Name : Gifu Anaerobic Broth (GAM)	Lot No : 0000573598
Report No.: 40001323092	Date of Release & Report : 2023-01-30	Expiry Date : 2027-12

other than that specified in the current HiMedia manual or product sheets. The results reported were obtained at the time of release.

This document has been produced electronically and is valid

Maya Sonavane

Microbiologist/Sr.Executive Microbiologist

Ujwala M. Kokate

Asst./Dy/QC Manager

Dr. Santosh Kaul

Dy/QA Manager

2023-01-30