

Liofilchem®

Certificate of Analysis

Page 1 of 1

| Product | Ref. | Batch | Expiration date |
|---|----------------------|-----------|-----------------|
| AD Fosfomicin 0.25-256 µg/mL | 77001 – 77061 | 090723004 | 2024.01.05 |

Microbiological Performance

Incubation: Ambient air, 16-20 h, 35 ± 2°C

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

| Control strains | Expected Results MIC range (mg/L) | Results MIC (mg/L) |
|--|--------------------------------------|-----------------------|
| <i>Staphylococcus aureus</i> ATCC® 29213 | 0.5–4* | 2 |
| <i>Enterococcus faecalis</i> ATCC® 29212 | 32–128 | 64 |
| <i>Escherichia coli</i> ATCC® 25922 | 0.5–2* | 1 |
| <i>Pseudomonas aeruginosa</i> ATCC® 27853 | 2–8* | 4 |

*Validated by EUCAST

Batch Release

Approved

Date

21.09.2023

Signature

Quality Control

(This document has been established electronically and is valid without signature)

The results reported were obtained at the time of release.