

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: PP3010 - Colorex Candida
Batch Number: 04168062
Expiry Date: 2024-01-02
Date Received: 2023-10-25
Date Tested: 2023-10-25
Date of Issue: 2023-10-30
Sample Condition: Satisfactory
Sample Number: 456972

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax: 01324 841314
Email: info@eolabs.com

The RGI is a calculation of the % growth on the test media compared with the growth on a control media. The test medium must achieve an RGI between 70-120% for non-selective media / $\geq 50\%$ for selective media.

Accredited Test Method: ED/SOP/008 Quantitative evaluation using spread inoculum technique (solid media)

| Productivity | RGI (%) | Colonial Appearance | Colonial Appearance Specification |
|---------------------------------|---------|---------------------|-----------------------------------|
| Candida tropicalis NCPF 3290 | 81 | Blue colonies | Blue colonies |
| Candida glabrata NCPF 3309 | 100 | Pink colonies | Pink colonies |
| Candida albicans NCPF 3179 | 72 | Green colonies | Green colonies |

| Physical | Result | Specifications | Accredited Test Method |
|---|----------|--------------------------|---|
| Sterility | Conforms | Within acceptable limits | ED/SOP/005 Visual check and growth assessment following incubation for 3 days at 15-25°C and 37°C |
| Sterility sampling is performed in accordance with ISO 2859-1:1999*. The inspection level is $\geq 0.4\%$ of the batch and the reject level ≤ 7 units depending on batch size. | | | |
| pH | 6.3 | 6.1 +/- 0.2 | ED/SOP/003 measurement by pH meter |
| Colour | Conforms | 2-8 - 6-8 Straw | ED/SOP/009 by visual observation. Range measured using Pantone® 4 colour process guide |

Certificate of Analysis

Client

E&O Laboratories Ltd
Burnhouse
Bonnybridge
Scotland
FK4 2HH



E&O Laboratories Ltd

Sample: PP3010 - Colorex Candida
Batch Number: 04168062
Expiry Date: 2024-01-02
Date Received: 2023-10-25
Date Tested: 2023-10-25
Date of Issue: 2023-10-30
Sample Condition: Satisfactory
Sample Number: 456972

Burnhouse, Bonnybridge
Scotland, FK4 2HH
Telephone: 01324 840404
Fax:01324 841314
Email: info@eolabs.com

| | | | |
|---------------|----------|------------|--|
| Fill Quantity | Conforms | 18-22ml | ED/SOP/054 by gravimetric determination |
| Moisture | Conforms | 95% +/- 3% | ED/SOP/053 with drying and gravimetric determination |

All of the results on this certificate of analysis relate only to the samples submitted.

Test specifications are based on ISO 11133:2014/Amd/:2020 and internal product specifications

*Sterility sampling is outwith scope of accreditation.



1894

Douglas Cameron
Technical Manager, E&O Laboratories Ltd