



Quality certificate

Biological indicators SIMICON ST/DA-V6

Biological indicators in compliance with ISO 11138-3 for steam sterilization at 121 °C/134 °C.

Organism: **Geobacillus stearothermophilus spores ATCC 7953**

Mean population (cfu): **1,7 × 10⁶**
determined by using agar surface method (TSA-Oxoid)

Carrier: **filter paper**

Packaging: **paper / foil**

Lot: **V210923**

Expiration date: **10.2025**

D_{121°C}-Value: **2,1 min**

z-value: **8,4 °C**

Survival time: **D-value x (log cfu - 2)**

Kill time: **D-value x (log cfu + 4)**

All data are determined at time of manufacturing.
Resistance data all determined according to ISO 11138 and are only reproducible when exposed and cultured under the exact conditions used to obtain results report here. Results may vary under different test conditions.

Incubation: **7 days at 56 °C ± 2 °C in CSB**

Storage:
at + 18 °C to + 25 °C and rel. humidity from 35 % to 70 %
Protect from light and sterilants.

Do not use after expiration date. Sterilize all cultures before disposal.

Date: 26.10.2023 Release: P.Barone

SIMICON GMBH • Sigmund-Riefler-Bogen 19 • D - 81829 München
Tel.: +49 / (0)89 / 67 33 66 0 • Fax: +49 / (0)89 / 67 33 66 22



Hygiene & Mikrobiologie

SIMICON GMBH
Sigmund-Riefler-Bogen 19
81829 München
Tel.: +49-89-673366-0
Fax: +49-89-673366-22

26.10.2023

Date

Stamp / Signature