

# QUALITY CONTROL CERTIFICATE



Product name : Legionella-GVPC-Selektivagar  
Abbreviation : GVPC  
Article code : 40-1602  
Filling Volume : 18 ml

Batch number : 16022101  
Storage temp. : 4-25 °C  
Expiration date : 29-02-2024  
*(for sealed enclosure as dd-mm-yyyy )*

## Typical Formula

Ingredients in gram per litre of nutrient medium \* :

Yeast extract	10	gr
Activated charcoal	2	gr
ACES buffer	10	gr
L-Cysteine	0.4	gr
Iron (III) pyrophosphate	0.25	gr
Alpha-Ketoglutarate	1	gr
Potassium hydroxide	2.8	gr
Agar	13	gr
Glycine	3	gr
Polymyxin B [IE]	80000	gr
Vancomycin [mg]	1	gr
Cycloheximide	0.08	gr

\* Adjusted as required to meet performance standards.

## Physical Properties

Test	Criterion	Result	Conforms
pH	6.8 ± 0.2	6.6	C
Color	black	black	C
Consistency	agar	agar	C
Gel Strength		Conforms	C
Appearance	Clear:No Partic.:No	Clear:No Part.:No	C

## Microbial Contamination

Incubation-condition	Result	Conforms
44-52 hours 30 ± 1 °C	No contamination	C

**Growth Properties on next page.**

# Quality Control Certificate

Product name : Legionella-GVPC-Selektivagar  
Abbreviation : GVPC  
Article code : 40-1602

Batch number : 16022101  
Storage temp. : 4-25 °C  
Expiration date : 29-02-2024

## Growth Properties

### • Incubation-condition : 116-124 hours 36 ±2 °C Aerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
Legionella pneumophila	ATCC 33152	06	grey-white	114	116	0.98	≥ 0.5	C
	WDCM 00107							
	NCTC 11192							
Legionella anisa	ATCC 35292	06	grey-white	74	85	0.87	≥ 0.5	C
	WDCM 00106							
	NCTC 11974							

### • Incubation-condition : 69-72 hours 37 ±1 °C Aerobic

Micro-organisms	Strain	Method*	Electivity**	CFU	CFU-ref	Result	Crit.	Conf.
Escherichia coli	ATCC 25922	07	cream	n.a.	32000	1	≤ 1	C
	WDCM 00013							
	NCTC 12241							
Enterococcus faecalis	ATCC 19433	07	n.a.	n.a.	16000	0	≤ 0	C
	WDCM 00009							
	NCTC 775							

\* Method 06 = TV 5.06 Selective enumeration agar ; Productivity of solid media ; quantitative method (acc. ISO 11133)  
Result: Productivity-Ratio (cfu test / cfu ref) ; inoculum 10<sup>2</sup> cfu

\* Method 07 = TV 5.07 Selective enumeration agar ; Selectivity of solid media ; qualitative method (acc. ISO 11133)  
Result: Growth ; 2 = good, 1 = slight, 0 = no growth ; inoculum 10<sup>4</sup>-10<sup>6</sup> cfu

\*\* Electivity denotes the properties of the colony and/or their effect on the morphological characteristics of the medium  
CFU denotes Colony Forming Units counted on batch. CFU-ref denotes Colony Forming Units counted on reference plate.  
n.a = Not Applicable. Crit.= Criterium C = Conforms  
Keep protected from light and at a constant temperature.  
The test results reported on this quality control certificate were obtained from a sample of the batch.

Xebios Diagnostics Group is certified according to DIN EN ISO 9001. Our microbiological quality control is carried out by our EN ISO/IEC 17025 accredited laboratory (RVA Registration no: RvA-L 614) at Tritium Microbiologie BV, Eindhoven.  
Tritium Microbiologie BV is part of Xebios Diagnostics Group.  
This laboratory certificate is generated electronically and valid without a signature.  
Results were obtained from laboratory report X20231102827.

Release Date :  
05-12-2023  
(dd-mm-yyyy)

Crit.version :7 Document ID : Q20231102827