

HiMedia Laboratories Private Limited

C-40, Road No.21Y, MIDC, Wagle Industrial Area,
Thane(W) - 400604 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity

Material Code : M1062	Material Name : *Dey-Engley Neutralizing Broth	Lot No : 0000615623
Report No.: 40001411369	Date of Release & Report : 2023-11-04	Expiry Date : 2027-10

Appearance

Light yellow to bluish grey homogeneous free flowing powder. Observed : Bluish grey

Colour and Clarity of prepared medium

Purple to reddish purple coloured, opalescent solution with slight precipitate forms in tubes.

Reaction

Reaction of 3.9% w/v aqueous solution at 25°C.

pH

pH Range :7.40-7.80 Observed : 7.77

Cultural Response

Cultural characteristics observed after an incubation at 35-37°C for 40- 48 hours.

Organism	Inoculum (CFU)	Growth
Cultural Response		
<i>Bacillus spizizenii</i> ATCC 6633 (WDCM 00003)	50-100	luxuriant
<i>Escherichia coli</i> ATCC 25922 (WDCM 00013)	50-100	luxuriant
<i>Pseudomonas aeruginosa</i> ATCC 27853 (WDCM 00025)	50-100	luxuriant
<i>Salmonella enterica</i> serovar Typhimurium ATCC 14028 (WDCM 00031)	50-100	luxuriant
<i>Staphylococcus aureus</i> ATCC 25923 (WDCM 00034)	50-100	luxuriant
<i>Escherichia coli</i> ATCC 8739 (WDCM 00012)	50-100	luxuriant
<i>Staphylococcus aureus</i> ATCC 6538 (WDCM 00032)	50-100	luxuriant

- . ATCC is a registered trade mark of the American Type Culture Collection
- . NCTC and National Collection of Type Culture are registered trade mark of the Health Protection Agency

- . All ISO 11133 : 2014/Amd.1:2018(E) control strains are included in the Quality parameter
- . HiMedia Laboratories Pvt Ltd is Certified for ISO 9001:2015, ISO 13485:2016 , WHO GMP

. Information for BSE/TSE Risk: The material was subjected to pH ≤ 7.0 and/or a temperature in excess of 75°C for no less than 2 hours during the manufacturing process. The bovine raw material for this product was collected entirely from Indian Origin animals in a licensed based establishment. The animals are inspected under a Govt. approved veterinarian's supervision and were

HiMedia Laboratories Private Limited

C-40, Road No.21Y, MIDC, Wagle Industrial Area,
Thane(W) - 400604 , Website : www.himedialabs.com,
Email : info@himedialabs.com

Certificate of Analysis, Quality and Conformity


Material Code : M1062	Material Name : *Dey-Engley Neutralizing Broth	Lot No : 0000615623
Report No.: 40001411369	Date of Release & Report : 2023-11-04	Expiry Date : 2027-10

apparently free from infectious and contagious diseases. BSE (Bovine Spongiform Encephalopathy)/ TSE (Transmissible Spongiform Encephalopathy) and dioxine are not known to exist in India. This material does not contain, nor is derived from the specific risks material as defined in The Maharashtra Animal Preservation Act Govt. of Maharashtra, India.


STATUS OF THE MATERIAL : APPROVED

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particulars use, other than that specified in the current HiMedia manual or product sheets. The results reported were obtained at the time of release.

This document has been produced electronically and is valid


Sheetal Shewale

**Microbiologist/Sr.Executive
Microbiologist**


Ujwala M. Kokate

Asst./Dy/QC Manager


Dr. Santosh Kaul

Dy/QA Manager

2023-11-04