

**HiMedia Laboratories Private Limited**

C-40, Road No.21Y, MIDC, Wagle Ind. Area, Thane(W) - 400604

Website : www.himedialabs.com, Email : info@himedialabs.com

**Certificate of Analysis , Quality and Conformity**

<b>Material Code : R096</b>	<b>Material Name :</b> Acid Phosphatase Reagent	<b>Lot No : 0000619948</b>
<b>AR No.: 40001420083</b>	<b>Date of Report : 2023-12-13</b>	<b>Exp. Date : 2024-05</b>

<b>TEST</b>	<b>SPECIFICATIONS</b>	<b>RESULTS</b>
<b><u>Appearance</u></b> 1 Appearance	Light yellow to brown colour solution	Complies
<b><u>Clarity</u></b> 1 Clarity	Hazy solution with precipitate. Note : Precipitate will not affect the performance criteria of the reagent.	Complies
<b><u>Cultural Response</u></b> 1 Cultural response	Add 0.2 - 0.3 ml of Acid Phosphatase reagent to a 24 - 48 hours old culture of the organism under investigation.	
<b><u>Clostridium perfringens ATCC 12924</u></b> 1 Acid Phosphatase	strong purple colour developed within 3-4min	Complies
<b><u>Clostridium perfringens ATCC 13124 (WDCM 00007)</u></b> 1 Acid Phosphatase	strong purple colour developed within 3-4 min	Complies
<b><u>Escherichia coli ATCC 25922 (WDCM 00013)</u></b> 1 Acid Phosphatase	No colour change	Complies

**STATUS OF THE MATERIAL : APPROVED**

This is to certify that this lot passes and it confirms to the above mentioned tests and specifications . The information given here is believed to be correct and accurate, however, both the information and products are offered without warranty for any particular use, other than that specified in the current HiMedia manual or product sheets.

**This document has been produced electronically and is valid**

**HiMedia Laboratories Private Limited**

C-40, Road No.21Y, MIDC, Wagle Ind. Area, Thane(W) - 400604

Website : www.himedialabs.com, Email : info@himedialabs.com

<b>Material Code : R096</b>	<b>Material Name :</b> Acid Phosphatase Reagent	<b>Lot No : 0000619948</b>
<b>AR No.: 40001420083</b>	<b>Date of Report : 2023-12-13</b>	<b>Exp. Date : 2024-05</b>

<b>TEST</b>	<b>SPECIFICATIONS</b>	<b>RESULTS</b>
 Mayuri A. Gorivale <b>Microbiologist/Sr.Executive Microbiologist</b>	 Ujwala M. Kokate <b>Asst./Dy/QC Manager</b>	 Dr. Santosh Kaul <b>Dy/QA Manager</b>