

Liofilchem®

Certificate of Analysis

Product	Batch	Production date	Expiration date
Cefiderocol FDC 0.016-256 µg/mL	022224042	22.02.2024	2025.02.21

Ref.

92067 – 920670 – 920671

Antimicrobial Susceptibility Testing

Tested according to current CLSI methodology

Control strains	Medium	Inoculum	Incubation	Expected Results MIC range (mg/L)	Results MIC(mg/L)
<i>Pseudomonas aeruginosa</i> ATCC® 27853	Mueller Hinton II Agar	0.5 McFarland in saline	35 ± 2°C, ambient 16-20 h	0.06–0.5	0.25

Batch Release

Approved

Date

28.02.2024

Signature

Quality Control
(This document has been
established electronically and
is valid without signature)

The results reported were obtained at the time of release.