

Liofilchem S.r.l.	CERTIFICATO CONTROLLO QUALITÀ	N°10127
	QUALITY CONTROL CERTIFICATE	Revisione 2 del 09.10.2013 Pag. 1 di 1

PRODOTTO / *PRODUCT* : **LEGIONELLA MWY AGAR**

LOTTO / *BATCH* : **050324037**

DATA PRODUZIONE/*PRODUCTION DATE*: **03.05.2024**

DATA SCADENZA / *EXP. DATE* : **2024.10.30**

TEST DI STERILITÀ / *STERILITY TEST*

20 ± 2°C (after 7 days)	–
37°C (after 7 days)	–

CARATTERISTICHE BIOLOGICHE / *BIOLOGICAL CHARACTERISTICS*

Tempo di incubazione / <i>Incubation time</i>	18h - 24 h *48h-72h
Temperatura / <i>Temperature</i>	37 °C ± 2 °C
Modalità di incubazione / <i>Procedure of incubation</i>	√ O ₂ w humidity CO ₂

Medium	LEGIONELLA MWY AGAR			
	Theoretical results		Result of conformity	
<i>Microorganism</i>	<i>Growth</i>	<i>Colour</i>	<i>Growth</i>	<i>Colour</i>
<i>Escherichia coli</i> ATCC 25922	±	Colorless	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<i>Staphylococcus aureus</i> ATCC 25923	±		<input checked="" type="checkbox"/>	
<i>Streptococcus faecalis</i> ATCC 19433	±		<input checked="" type="checkbox"/>	
* <i>Legionella pneumophila subsp. fraseri</i> ATCC 33156	+	Gray	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<i>pH at 25 °C</i>	7.0 ± 0.2		7.1	
(+) = <i>Crescita / Growth</i> (-) = <i>Inibizione / Inhibition</i> (±) = <i>Inibizione parziale/Partial inhibition</i>				

LOTTO/BATCH : Idoneo / Approved

Non Idoneo / Not approved

DATA / *DATE* 10.05.2024

Responsabile Controllo Qualità
Quality Control Manager
(This document has been
established electronically and is valid without
signature)