


Certificate of Analysis: Lyophilized Microorganism Specification and Performance Upon Release

<p>SPECIFICATIONS:</p> <p>Product Name: Staphylococcus aureus subsp. aureus</p> <p>Catalog Number: 0485</p> <p>Lot Number: 485-1233**</p> <p>Reference Number: ATCC® 6538™*</p> <p>Passage from Reference: 3</p> <p>(7) Mean Assay Value (MAV): 62 CFU per 0.1 ml</p> <p>Expiration Date: 2025/11/30</p>	<p>RELEASE INFORMATION:</p> <p>Quality Control Technologist: Madison C Springer</p> <p>Release Date: 2023/12/04</p>
--	--

Performance	
<p>Macroscopic Features:</p> <p>Medium to large, convex, circular, glistening, smooth, creamy, opaque, beta hemolytic - both light gold and darker gold colonies may be present. A second colony type maybe present a white, circular, entire, low convex, and beta hemolytic.</p> <p>Microscopic Features:</p> <p>N/A</p>	<p>Medium:</p> <p>SBAP</p> <p>Method:</p> <p>N/A</p>

ID System: MALDI-TOF (1)

See attached ID System results document.



Amanda Kuperus
Director of Quality Control
AUTHORIZED SIGNATURE

**Disclaimer: The last digit(s) of the lot number appearing on the product label and packing slip are merely a packaging event number. The lot number displayed on this certificate is the actual base lot number.

Refer to the enclosed product insert for instructions, intended use and hazard/safety information.

Individual products are traceable to a recognized culture collection.

(1) These tests are accredited to ISO/IEC 17025.



(*) The ATCC Licensed Derivative Emblem, the ATCC Licensed Derivative word mark and the ATCC catalog marks are trademarks of ATCC. Microbiologics, Inc. is licensed to use these trademarks and to sell products derived from ATCC® cultures.



(7) The Mean Assay Value (MAV) stated above may deviate from the end-user's MAV based on variables inherent to each laboratory environment, such as methods, media type, equipment, pipettes, and individual technician technique.

Bruker Daltonik MALDI Biotyper Classification Results



Meaning of Score Values

Range	Interpretation	Symbols	Color
2.00 – 3.00	High-confidence identification	(+++)	Green
1.70 – 1.99	Low-confidence identification	(+)	Yellow
0.00 – 1.69	No Organism Identification Possible	(-)	Red

Meaning of Consistency Categories (A - C)

Category	Interpretation
(A)	High consistency: The best match is a high-confidence identification. The second-best match is (1) a highconfidence identification in which the species is identical to the best match, (2) a low-confidence identification in which the species or genus is identical to the best match, or (3) a non-identification.
(B)	Low consistency: The requirements for high consistency are not met. The best match is a high- or low-confidence identification. The second-best match is (1) a high- or low-confidence identification in which genus is identical to the best match or (2) a non-identification.
(C)	No consistency: The requirements for high or low consistency are not met.

Run Creation Date/Time: 2023-11-29T10:58:05.173 JAL

Applied MSP Library(ies): BDAL, Mycobacteria Library (bead method), Filamentous Fungi Library

Sample Name	Sample ID	Organism (best match)	Score Value
A11 (+++) (A)	485-1233	Staphylococcus aureus	2.10

Comments:

N/A



Statistical Analysis Certificate

Microorganism Name: Staphylococcus aureus subsp. aureus

Reference #: ATCC® 6538™*

Catalog #: 0485

Lot #: 485-1233**

Expiration Date: 2025/11/30

(7) Mean Assay Value (MAV): 62 CFU per 0.1 ml

Standard Deviation: 8.2E+00

Coefficient of Variation: 13%

99% Confidence Interval of 6.0E+01 to 6.5E+01 CFU

95% Confidence Interval of 6.0E+01 to 6.5E+01 CFU

Method used to determine Mean Assay Value: Spiral Plate Method

Medium Employed: TSA

Incubation Time and Temp: 24 hrs at 34-38 degrees C

A handwritten signature in black ink, appearing to read "Amanda Kuperus". The signature is fluid and cursive, written over a white background.

Amanda Kuperus

Director of Quality Control

AUTHORIZED SIGNATURE

(7) The Mean Assay Value (MAV) stated above may deviate from the end-user's MAV based on variables inherent to each laboratory environment, such as methods, media type, equipment, pipettes, and individual technician technique.

The last digit(s) of the lot number appearing on the product label and packing slip are merely a packaging event number. The lot number displayed on this certificate is the actual base lot number. The information included in this Statistical Analysis Certificate is strictly based on the product's lot number. A product lot number may be assigned to multiple packaging configurations. As a result, this certificate only lists the lot number and does not include a product description.

© 2012 Microbiologics, Inc. All Rights Reserved. 200 Cooper Avenue North Saint Cloud, MN 56303