



PROVEN DIAGNOSTIC PERFORMANCE

CERTIFICATE OF ANALYSIS

Product: LACTOFERRIN SCAN™ Catalog: T5009
Lot: 0124164 Expiration: 2024/12/01

Date of Manufacture: 2023/09/20

Biological components and specifications:
All components meet individual release criteria.

Table with 4 columns: Description, Product, Lot #, Exp. Date. Lists various components like Conjugate, Diluent, Stop Solution, Standards, Control, Substrate, Wash Solution, and Plate Sealer.

Kit functionality:

The LACTOFERRIN SCAN™ test has been tested and released to meet the criteria detailed below:

Positive Control = 10 ± 3 µg/mL
Negative Control (1X Diluent) OD450 < 0.100 OD450/620 < 0.060
R² ≥ 0.98

Ingredients of animal origin are sourced from countries/areas with no-risk or low-risk status for bovine spongiform encephalopathy (BSE).

This product has been manufactured in accordance with the requirements of the TECHLAB, Inc. Quality Management System. TECHLAB® performs quality control and release testing of all in vitro diagnostic test kits manufactured and released for sale.

Logan Hollyfield 11/2/2024
Quality Assurance Representative/Date

Revision: 05-26-2022