

CERTIFICATE OF ANALYSIS

Product:

H. PYLORI QUIK CHEK™

Catalog:

30925

Lot:

0424025

Expiration:

2026/02/01

Date of Manufacture: 2024/03/29

Biological components and specifications: All components meet individual release criteria.

Description	Product #	Lot#	Exp. Date
Conjugate	5050C	0324124	2026/03/01
Diluent	5050D	0324125	2026/03/01
Membrane Device	5050MD	0324126	2026/03/01
Positive Control	5050P	0324127	2026/03/01
Membrane Substrate	MS2	0224135	2027/10/17
Wash Buffer	MW2	0224136	2026/02/01
T (D:			
Transfer Pipettes	93-219-00		
Wooden Applicator Sticks	93-110-01		
Product Insert	91-30925-01-TL		
Instructional Card	92-050-02		

Kit functionality:

The H. PYLORI QUIK CHEK™ test has been tested and released to meet the criteria detailed below:

Positive Control = Visible Control Line and Visible Test Line

Negative Control (Diluent) = Visible Control Line and No Visible Test Line

Ingredients of animal origin are sourced from countries/areas with no-risk or low-risk status for bovine spongiform encephalopathy (BSE).

This product has been manufactured in accordance with the requirements of the TECHLAB, Inc. Quality Management System. TECHLAB performs quality control and release testing of all *in vitro* diagnostic test kits manufactured and released for sale. The results of this testing are documented per the TECHLAB Quality Management System. A Certificate of Analysis is issued for each lot of material tested and passed against required Quality Control release specifications.

Bethony Stike 4/17/2024

Quality Assurance Representative/Date

Revision: 01-13-2023