

<b>Liofilchem S.r.l.</b>	CERTIFICATO CONTROLLO QUALITÀ	N°10127
	<i>QUALITY CONTROL CERTIFICATE</i>	Revisione 2 del 09.10.2013 Pag. 1 di 1

PRODOTTO / *PRODUCT* : **LEGIONELLA MWY AGAR**

LOTTO / *BATCH* : **070524023**

DATA PRODUZIONE/*PRODUCTION DATE*: **05.07.2024**

DATA SCADENZA / *EXP. DATE* : **2025.01.01**

**TEST DI STERILITÀ / *STERILITY TEST***

20 ± 2°C (after 7 days)	–
37°C (after 7 days)	–

**CARATTERISTICHE BIOLOGICHE / *BIOLOGICAL CHARACTERISTICS***

Tempo di incubazione / <i>Incubation time</i>	18h - 24 h *48h-72h
Temperatura / <i>Temperature</i>	37 °C ± 2 °C
Modalità di incubazione / <i>Procedure of incubation</i>	√ O <sub>2</sub> w humidity      CO <sub>2</sub>

<b>Medium</b>	<b>LEGIONELLA MWY AGAR</b>			
	<b>Theoretical results</b>		<b>Result of conformity</b>	
<i>Microorganism</i>	<i>Growth</i>	<i>Colour</i>	<i>Growth</i>	<i>Colour</i>
<i>Escherichia coli</i> ATCC 25922	±	Colorless	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<i>Staphylococcus aureus</i> ATCC 25923	±		<input checked="" type="checkbox"/>	
<i>Streptococcus faecalis</i> ATCC 19433	±		<input checked="" type="checkbox"/>	
* <i>Legionella pneumophila subsp. fraseri</i> ATCC 33156	+	Gray	<input checked="" type="checkbox"/>	<input checked="" type="checkbox"/>
<i>pH at 25 °C</i>	7.0 ± 0.2		6.8	
(+) = <i>Crescita / Growth</i> (–) = <i>Inibizione / Inhibition</i> (±) = <i>Inibizione parziale/Partial inhibition</i>				

LOTTO/BATCH :  Idoneo / Approved

Non Idoneo / Not approved

DATA / *DATE* 12.07.2024

Responsabile Controllo Qualità  
*Quality Control Manager*  
(This document has been  
established electronically and is valid without  
signature)